

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर

GOVERNMENT MEDICAL COLLEGE & HOSPITAL, NAGPUR

FAX NO. 0712-274448

Tele. No. Office:-2743588

No.Hosp/PS/Qt/Enq/

708

/2022

Dt. 16/01/23

Quotation

To,

M/s.....

.....

.....

.....

.....

Subject :- Submission of Quotation for the Supply of Instrument/Articles/Materials to the  
Govt. Medical College & Hospital, Nagpur

The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the Govt. Medical College & Hospital, Nagpur on the following terms & conditions.

- 1) The prices quoted should be free delivery to College & Hospital premises for local dealers and F.O.R. Nagpur for outsiders.
- 2) The Price quoted by inclusive of all taxes, duties if payable like customs, excise, CST, ST, and GST. The break-up of the taxes should also be shown separately where necessary. The sales tax & registration number should be quoted in your letter. Exemption of taxes, if on A.F. Forms etc. be separately stated.
- 3) The serial number of the items should not be changed while quoting rates. You may drop the item if not interested.
- 4) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 5) The quotation submitted will be valid for a period of One (01) year only from the date of acceptance.
- 6) Delivery period should be stated specifically the like ready stock or two weeks or four weeks etc and should be firm and supply of stores if ordered should be made available within the stipulated period. Failures are liable for dislist from further enquiries.
- 7) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover. Unsealed quotations which is not properly sealed will not be accepted. The last or receiving date of the quotation is Dt. 23 Jan 2023 upto 3.00 p.m.
- 8) Supply of stores should be made in one installment unless otherwise ordered place meal supply will not be accepted. Condition of advance payment through Bank, partpayment etc. will not be accepted.
- 9) The price quoted is inclusive of all taxes, duties if payable like Customs/ Excise/CST/ST the break up of the taxes should also be shown separately where necessary. The sales tax and registration number should be quoted in your letter. Exemption of taxes on 'AF'/ Octroi exemption form etc be separately stated.

(Contd. On ...2..)

...2...

10) Quotations if asked with samples if not accompanied with sample will be liable for rejection even they are lowest. Samples should be sent with labels attached quoting our ref. no of enquiry and item no etc.

11) The Dean, Govt. Medical College & Hospital, Nagpur does not pledge himself to accept the lowest or any quotations and reserves to himself right of acceptance or rejection any of quotations which does not suit to his requirements.

12) Very Important :- In case you are not interested in quoting your rates; in reply to this enquiry a line in reply is must; failing are liable for dislisting their names for further enquiries from our list and no further request in this matter will be entertained.

13) Literature, Instruction manual, working leaflet etc may also be sent with quotation.

14) Authorization certificate, Brochure, Catalogue, User List may also submit alongwith quotation is mandatory.

15) Technical Specifications - Compliance Chart with variation/daviation

  
DEAN


Govt. Medical College & Hospital, Nagpur


Details as below


Sr. No.	Name of the Items	Rates including GST
1	Anesthesia Machine (Specification attached)	

## Technical Specification Anaesthesia Boyle's Machine

Sr. No.	Specification of Anaesthesia Machine with Tec- 7 vaporizer- 1& with close circuit - 1	Specification of Anaesthesia Machine with Tec - 7 vaporizer - 1 & with close circuit - 1
1	Anaesthesia Machine : Flow Meter- 2 or 3 tube rota meter	Anaesthesia Machine : flow Meter - 2or 3 tube rota meter - Provided
2	Material : MS/ fibre body with 4 wheels ( two wheels with brakes)	It has material : MS/ fibre body with 4 wheels (two wheels with brakes)- Provided
3	Design: As per ISO norm with preset regulator valves for safety of pt & machine	It has Design: As per ISO norm with preset regulator valves for safety of pt & machine - Provided
4	Audiovisual alarm for gas failure	It has Audiovisual alarm for gas failure - Provided
5	FGF outlet for open / close circuit	It has FGF outlet for open/ close circuit - Provided
6	Provision for oxygen supply - Two part [ One for oxygen cylinder & one for central supply	It has Provision for oxygen supply - two part [ one for oxygen cylinder & one for central supply ] - Provided
7	Provision for Nitrous Oxide supply - Two part [ one for N2O cylinder & one for central supply]	It has Provision for nitrous oxide supply - Two part [one for N2O cylinder & one for central supply]- provided
8	Emergency flush valve & emergency high pressure relief valve	It has emergency flush valve & emergency high pressure relief valve- provided
9	Selectatec bar for two vaporizer	It has selectatec bar for two vaporizer provided
10	Hypoxic Guard & master slave safety system	It has Hypoxic Guard & Master Slave safety system Provided
11	Close circuit circle system with reusable patient circuit	Close circuit circle system with reusable patient circuit - Provided
12	Bain's circuit	It has Bain's circuit - provided
13	JR Paediatric circuit	It has JR paediatric circuit - provided
14	Scope of Supply	Scope of supply provided
15	Anaesthesia Machine Unit - 1no	It has Anaesthesia Machine Unit - 1no provided
16	Vaporizer: Sevoflurane / Isoflurane Tec 7 - 1 no	It has Vaporizer : Sevoflurane / Isoflurane (Tec- 7) 1-no provided
17	Close circuit reusable patient circuit (Adult)	It has Close circuit reusable patient circuit (Adult) 1No. Provided
18	Bain's circuit	It has Bain's circuit 1no. Provided
19	JR Paediatric circuit	It has JR Paediatric circuit 1no. Provided
20	High pressure Oxygen tubing (White colour) (8-10 Mtr length with pin Index yoke block for oxygen ( At one end) & Oxygen probe front loaded Compatible ( at other end ) - 1 no.	High pressure Oxygen tubing ( white colour) (8-10 mtr length with pin index yoke block for oxygen (At one end ) & Oxygen probe front loaded Compatible ( at other end) - 1NO Provided

  
 सहाय्यी प्राध्यापक  
 वसिरीकरण शास्त्र विभाग  
 सा. वै. महा. नागपूर

  
 सहाय्यी प्राध्यापक  
 वसिरीकरण शास्त्र विभाग  
 सा. वै. महा. नागपूर

  
 20/12/22  
 Professor & Head  
 Dept. of Anaesthesia

**Technical Specification  
Anaesthesia Boyle's Machine**

21	High pressure Nitrous oxide tubing (White colour) (8-10 mtr length with pin Index yoke block for oxygen ( At one end) & Oxygen probe front loaded Compatible ( at other end ) - 1 no.	High pressure Nitrous oxide tubing ( white colour) (8-10 mtr length with pin index yoke block for oxygen (At one end ) & Oxygen probe front loaded Compatible ( at other end) - 1NO Provided
22	Any other accessories as per anaesthesia machine specs	Any other accessories as per anaesthesia machine specs Provided.

सहसंचालक  
वैद्यकीय शिक्षण व संशोधन, मुंबई

संचालक  
वैद्यकीय शिक्षण व संशोधन, मुंबई

*20/12/22*  
*K. Lindsey*  
सहयोगी प्राध्यापक  
वैद्यकीय शास्त्र विभाग  
शा. वै. महा. नागपूर

**Professor & Head  
Dept. of Anesthesia  
Govt. Medical College, Nagpur**

*M.L.L.*  
*21/12/22*  
सहयोगी प्राध्यापक  
वैद्यकीय शास्त्र विभाग  
शा. वै. महा. नागपूर