

# शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर

GOVERNMENT MEDICAL COLLEGE & HOSPITAL, NAGPUR

Tele. No. Office:-2743588

FAX NO. 0712-274448

No.Hosp/PS/Qt/Enq/

/2022

Dt.

Quotation

1320

24/01/2023

To,

M/s.....

.....

.....

.....

Subject :- Submission of Quotation for the Supply of Instrument/Articles/Materials to the  
Govt. Medical College & Hospital, Nagpur

The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the Govt. Medical College & Hospital, Nagpur on the following terms & conditions.

- 1) The prices quoted should be free delivery to College & Hospital premises for local dealers and F.O.R. Nagpur for outsiders.
- 2) The Price quoted by inclusive of all taxes, duties if payable like customs, excise, CST, ST, and GST. The break-up of the taxes should also be shown separately where necessary. The sales tax & registration number should be quoted in your letter. Exemption of taxes, if on A.F. Forms etc. be separately stated.
- 3) The serial number of the items should not be changed while quoting rates. You may drop the item if not interested.
- 4) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 5) The quotation submitted will be valid for a period of One (01) year only from the date of acceptance.
- 6) Delivery period should be stated specifically the like ready stock or two weeks or four weeks etc and should be firm and supply of stores if ordered should be made available within the stipulated period. Failures are liable for dislist from further enquiries.
- 7) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover. Unsealed quotations which is not properly sealed will not be accepted. The last or receiving date of the quotation is Dt. 02 Feb 2023 upto 3.00 p.m.
- 8) Supply of stores should be made in one installment unless otherwise ordered place meal supply will not be accepted. Condition of advance payment through Bank, partpayment etc. will not be accepted.
- 9) The price quoted is inclusive of all taxes, duties if payable like Customs/ Excise/CST/ST the break up of the taxes should also be shown separately where necessary. The sales tax and registration number should be quoted in your letter. Exemption of taxes on 'AF' Octroi exemption form etc be separately stated.

(Contd. On ...2..)

...2...

10) Quotations if asked with samples if not accompanied with sample will be liable for rejection even they are lowest. Samples should be sent with a label attached quoting our ref. no of enquiry and item no etc.

11) The Dean, Govt. Medical College & Hospital, Nagpur does not pledge himself to accept the lowest or any quotations and reserves to himself right of acceptance or rejection any of quotations which does not suit to his requirements.

12) Very Important :- In case you are not interested in quoting your rates; in reply to this enquiry a line in reply is must; failing are liable for dislisting their names for further enquiries from our list and no further request in this matter will be entertained.

13) Literature, Instruction manual, working leaflet etc may also be sent with quotation.

14) Authorization certificate, Brochure, Catalogue, User List may also submit alongwith quotation is mandatory.

15) Technical Specifications - Compliance Chart with variation/davation



DEAN

Govt. Medical College & Hospital, Nagpur

Details as below

Sr. No.	Name of the Items	Rates including GST
1	Autoclave Machine (As per specification attached)	

**Name of Equipment:- Autoclave-Top Loading**

**Specifications for Autoclave-Top Loading**

- Should be electrically operated, vertical type top loading with vertical type with locking lid
- It should be triple wall constructed
- locking of lid by heat resistant handle/nuts/radial locking system with good quality rubber gasket
- body made up of good quality material (stainless steel / aluminum)
- It should be provided with built- in safety valve, pressure gauge, pressure release valve and water level indicator / Marker
- Size- 25 liters
- It should have heating element (coil) inside
- pressure should be adjustable from 5 psi to 25 psi with automatic pressure control switch
- Temperature  $121^{\circ}\text{C}$ .
- It should be supplied with good quality outer stand
- Supplied with sterilization drum which well fits in autoclave, cord and plug.



*Apewo*

सहायक प्राध्यापक  
वैद्यशास्त्र विभाग

महाराष्ट्र वैद्यकीय महाविद्यालय व रुग्णालय  
नागपूर - ४४०००३

*Dr. S. S. S.*