

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर

GOVERNMENT MEDICAL COLLEGE & HOSPITAL, NAGPUR

Tele. No. Office:-2743588

FAX NO. 0712-274448

No.Hosp/PS/QT/Enq / /2023

Dt. 21 / 9 / 2023

Quotation  
UPC

17883

To,

M/s.....

Subject: - Submission of Quotation for the Supply of Instrument/Articles/Materials to the Govt. Medical College & Hospital, Nagpur.

The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the Govt. Medical College & Hospital, Nagpur on the following terms & conditions.

- 1) The prices quoted should be for delivery to College & Hospital premises for local dealers and F.O.R. Nagpur for outsiders.
- 2) The Price quoted by inclusive of all taxes, duties if payable like customs, excise, CST, ST, and GST. The break-up of the taxes should also be shown separately where ever necessary. The sales tax & registration number should be quoted in your letter. Exemption of taxes, if on A.F. Forms etc.be separately stated.
- 3) **The serial number of the items should not be changed while quoting rates.** You may drop the item if not interested.
- 4) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 5) The quotation submitted will be valid for a period of One (01) year only from the date of Acceptance.
- 6) Delivery period should be stated specifically , like ready stock or two weeks or four weeks etc and should be made available to this store within the stipulated period. Failures are liable for delist from further enquiries.
- 7) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover. Unsealed quotations which are not properly sealed will not be accepted. The last or receiving date of the quotation is **Dt. 27 Sept 2023** upto 3.00 p.m.
- 8) Supply of stores should be made in one instalment unless ordered otherwise. Piecemeal supply will not be accepted. Condition of advance payment through Bank, part payment etc. will not be accepted.
- 9) The price quoted is inclusive of all taxes, duties if payable like Customs/ Excise/CST/ST the breakup of the taxes should also be shown separately where necessary. The sales tax and registration number should be quoted in your letter. Exemption of taxes on 'AF/ Octroi exemption from etc. is separately stated.

(Contd. On ...2..)

...2...



- 10) Quotations if asked with samples should be accompanied if not accompanied with sample it will be liable for rejection even they are lowest. Samples should be sent with labels attached quoting our ref.no of enquiry and item no etc.
- 11) The Dean, Govt. Medical College & Hospital, Nagpur does not pledge himself to accept the lowest or any quotations and reserves the right of acceptance or rejection of any quotations which are not as per the specification or requirement .
- 12) Supplier must quote the rates of items only those which he/she will display/demonstrate and supply. If the good supplied are different from the sample displayed or checked , order will be cancelled & no payment will be done.
- 13) Literature, Instruction manual, working leaflet etc may also be sent with quotation.
- 14) Registration/Authorization Certificate, Sale tax, IT Return Certificate, PAN card, GST no Compliance chart ,brochure ,Catalogue, User List submission (wherever applicable) along with quotation is mandatory.
- 15) Technical Specifications - Compliance Chart with variation/deviation
- 16) While submitting the bills of good supplied, supplier must calculate unit price and final unit price should reflect the same rate as per order given by the Institute. GST & other taxes should not be calculated on total prices.
- 17) After Opening the technical BD qualified Bidder will have to give demonstration of equipment as well as show their quoted items to HOD for technical approval as a Technical Committee otherwise Bidder names not to be included in commercial bid.

Details as Below:

| Sr. No                                      | Name of Item with  | Qty   | Rate Incl.GST |
|---|--|-------|---------------|
| <b>CHEMICAL &amp; BIOLOGICAL INDICATORS</b> |  |       |               |
| 1   | <b>Class 1 Indicators STEAM - 3 Line Double Adhesive STEAM label</b> - color change after sterlization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 9001, ISO 13485.   | 01Nos |               |
| 2   | <b>Class 1 Indicators ETO - 3 Line Double Adhesive ETO label</b> - color change after sterlization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 13485  | 01Nos |               |
| 3   | <b>Class 2 Indicators STEAM- Bowie Dick test Hollow PCD 4</b> Metres Lumen length, 2 mm diameter, SS Device, Color Change from Yellow to Black in 6 separate block indicators. Non-Toxic, Lead Free, EN 285, EN ISO 11140 - Part 1:2009 and Part 4:2007 adherence. ISO 9001, ISO 13485. <b>1 PCD - Will be given Free in consignment</b> | 01Nos |               |
| 4   | <b>Class 5 Indicators - Biological Indicators - Steam Self Contained</b> $1 \times 10^6$ Population with Certificate of Analysis of each Batch. ISO 11138-1, ISO 11138-3, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value not lower than 1.5 min (121 Degrees C)  | 01Nos |               |
| 5   | <b>Class 5 Indicators - Biological Indicators - ETO Self Contained</b> $1 \times 10^6$ Population with Certificate of Analysis of each Batch. ISO 11138-1, ISO 11138-3, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value not lower than 2.5 min (600 mg/lEO))  | 01Nos |               |



|  |  |       |  |
|--|--|-------|--|
| 6  | <b>Class 6 Indicators STEAM-</b> Emulating Indicator " Not Sterile"<br>Mentioned with clear color change from Yellow to black only, With Ref Color Change, With Self-Adhesive Sticking, Double Laminated, Non-Toxic, Lead-Free, Metal Free. ISO 11140-1 Type 6 and Lot Number should be mentioned on Each Indicator for Lot Tracking. ISO 9001, ISO 13485  | 01Nos |  |
| 7  | <b>Class 5 Indicator ETO- Batch Monitoring System (BMS) -</b><br>Lot Control Test with Helix PCD of 1.5 Metres Length, 2 mm lumen dia, Teflon coated. 4 Block Indicator Color Change from light Blue to Green. 100% Toxic and Metal Free Substances, EN 285, EN 867-5, ISO 11140-1, ISO 11140-4, ISO 11140-5 Standards, ISO 13485 Certified  | 01Nos |  |
| <b>TVYEK Packing Sterilisation Reels (For Plasma H2O2) ISO 9001, ISO 13485, CE, EN 868 with Plasma Indicator, 1073B Tyvek</b>  |  |       |  |
| 8  | Tyvek Sterilization Roll 50mm x 70 M   | 01Nos |  |
| 9  | Tyvek Sterilization Roll 75mm x 70 M   | 01Nos |  |
| 10   | Tyvek Sterilization Roll 150mm x 70 M  | 01Nos |  |
| 11   | Tyvek Sterilization Roll 200mm x 70 M  | 01Nos |  |
| 12   | Tyvek Sterilization Roll 250mm x 70 M  | 01Nos |  |
| 13   | Tyvek Sterilization Roll 300mm x 70 M  | 01Nos |  |
| 14   | Tyvek Sterilization Roll 350mm x 70 M  | 01Nos |  |
| <b>Disposable SMS Wrapping Paper - US FDA Certified, CE Certified, 96.2% Bacterial Filtration Efficiency (Minimum) with Test Reports, Validation Reports, Certificates, MSDS</b> |  |       |  |
| 15   | Kinguard - 61 cm x 61 cm   | 01Nos |  |
| 16   | Kinguard - 76 cm x 76 cm   | 01Nos |  |
| 17   | Kinguard - 101 cm x 101 cm   | 01Nos |  |
| 18   | Kinguard - 121 cm x 121 cm   | 01Nos |  |
| <b>CSSD Disinfectant Solutions</b>   |  |       |  |
| 19   | <b>Multi-Enzyme Solution</b><br>12% v/v Protease, Lipase, Amylase with Non-Ionic Surfactant, PH Neutral, Corrosion inhibitors, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485   | 01Nos |  |
| 20   | <b>Rust &amp; Stain Remover</b><br>Inorganic Acid (> 30% Phosphoric Acid), Non Ionic Surfactant (<5%), Non-flammable, No harmful VOC's, GMP Certified, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485   | 01Nos |  |
| 21   | <b>High-Level Disinfectant</b><br>Aldehyde Free - 3% w/v [poly(hexamethylene biguanide)hydrochloride] (PHMB), 10% w/v Didecyl dimethyl ammonium chloride (DDAC) with Sporicidal and Mycobactericidal activity, Colourless, odourless, known action against nosocomical infections, biofilms, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485 | 01Nos |  |
| 22   | <b>Linen Disinfectant</b><br>5% w/v Didecyl Dimethyl Ammonium Chloride (DADC) 4th Gen QAC, bioburden tolerant, Contains fabric softener, no decoloring effect, Bactericidal, fungicidal, virucidal With Validation Reports, MSDS, CoA, Dilution Charts, Training on  | 01Nos |  |

|    |                         |       |  |
|----|-------------------------|-------|--|
| 41 | Humid chip              | 01Nos |  |
| 42 | ETO Bags                | 01Nos |  |
| 43 | Printer Roll            | 01Nos |  |
| 44 | Plasma Sterrad Cassette | 01Nos |  |
| 45 | Gasket                  | 01Nos |  |
| 46 | Bacterial Filter        | 01Nos |  |
| 47 | Lubricant               | 01Nos |  |

(As per specifications)

NOTE: - Price of items must be quoted inclusive of all taxes.



DEAN  
Govt. Medical College & Hospital,  
Nagpur