

महाराष्ट्र शासन  
शासकीय वैद्यकीय महाविद्यालय व अति विशेषोपचार रुग्णालय, नागपूर.  
Govt. Medical College & Super Speciality Hospital, Nagpur.

Phone No.0712-2701541

Fax No.0712-2746682

E-mail - gmcssh95@gmail.com

Ref.No.GMC&amp;SSH/PS/MJPJAY/CVTS OT/Perfusion (Heart Valve, Cannula &amp; Catheter)/QT/6178 /2023 Date:- 25/11/2023

UPC  
QUOTATION

To,  
M/s. ....  
.....  
.....

Sub :- Supply of Instrument/Equipments/Materials of the Govt. Medical College & Super Speciality Hospital, Nagpur. **(Quotation of Surgical Items for CVTS OT Perfusion Department) (List attached)**

Dear Sir,

The undersigned is invited sealed quotations for the store enlisted below/ As per enclosed statement, for the use of Govt. Medical College & Super Speciality Hospital, Nagpur on the following terms & conditions.

1. The prices quoted should be for free delivery of the ordered stores to Govt. Medical College & Super Speciality Hospital, Nagpur premises F.O.R. Destination.
2. The prices quoted should be inclusive of taxes of duties. If any payable like customs excise, CST, ST and G.T. the breakup of the taxes should also be shown separately wherever necessary. The Sales Tax Registration number should be quoted in your letters. Exemption of taxes, if on A.F. Forms also be separately stated.
3. The serial number of the items should not be changed while quotation rates. You may drop the item if not interested.
4. Rates should be quoted strictly for the item specified in the list & for standard quantity of goods. In case of alternate Officer, the detailed specification, name of manufacturer or make etc. must be stated. Specification other than specified in the schedule may be liable for rejection even through lowest.
5. The quotation submitted will be valid for a period of one year only from the date of acceptance.
6. Delivery period be stated specifically like ready Stock/Two weeks/Four weeks etc. & should be from supply of stores if order should be made within the stipulated period mentioned there in. Failures will be liable for delisting from further offers.
7. The quotation received after the due date will not be accepted. The quotations must be submitted in a sealed cover, Unsealed quotations will not be accepted. **The last date of receiving the quotation is 12/12/2023 by 4:00 P.M.**

8. Supply of ordered stores should be made in one installment unless otherwise ordered piecemeal supply will not be accepted. Payment will be made within 4 to 8 weeks after receive of full quantity with bills in quadruplicate & only after the satisfactory report of working etc. of the stores. Condition of advance payment through Bank. Etc. or part payment will not be accepted.
9. Quotations if asked with samples & if not accompanying the same will be liable for rejection even they are lowest samples should be sent with a tables attached quoting our ref. No. of Inquiry & Item No. etc.
10. The Govt. Medical College & Super Speciality Hospital, Nagpur do not pledge himself to accept the lowest or any quotation and reserve to right of acceptance of any quotation which suits to his requirements.
11. Incase you are not interested in quoting your rates in reply to his inquiry a line in reply is must. Failing which are liable for desisting their names for further enquires from our margin list and No further request in this matter will be entertained.
12. Literature, instructions manuals showing specifications, working etc. may also be sent with quotations.
13. Copy of Firm Registration Certificate/ST/CST/Registration No. GST No etc. should be mentioned in the quotation while submitting the same otherwise it will treat as invalid.
14. Bidder must show Quoted surgical items to concern HOD within 3 - 5 days after stipulated period (i.e. after final submission of quotation date) for technical approval. If you failed to provide the sample within stipulated period you will be considered disqualified and will not be considered for further quotation process.
15. Technical Bid & Commercial Bid Must be Submitted Separately in sealed cover, at the Dean Office GMC, Nagpur in a special quotation box of Super Speciality Hospital.



Dean,

Government Medical College &  
Super Speciality Hospital,  
Nagpur.

**महाराष्ट्र शासन**  
शासकीय वैद्यकीय महाविद्यालय व अति विशेषोपचार रुग्णालय, नागपूर.  
Govt. Medical College & Super Speciality Hospital, Nagpur.

Phone No.0712-2701511

Fax No.0712-2746682

E-mail - gmcssh95@gmail.com

Ref.No.GMC&amp;SSH/PS/MJPJAY/CVTS OT/Perfusion (Heart Valve, Cannula &amp; Catheter)/QT/ 6178 /2023


Date:-25/11/2023

### Surgical Items Requirement Quotation For CVTS OT Perfusion Department

(It is Compulsory to provide sample for Technical Approval)

Sr. No	Name Of Item	Specifition	Name Of Company/Brand/Make	Rate Per Pcs. Inclusive All Taxes With GST
1	Heart Valve Mechanical Bi- Leaflet Aortic 17mm/19mm/21mm/22mm/23mm/25mm/All Sizes	<ul style="list-style-type: none"> <li>• Pyrolytic carbon Mechanical Heart Valve for Aortic Heart position.</li> <li>• Bi-Leaflet U.S FDA &amp; CE approved.</li> <li>• Radio opaque for improved visualization during X-ray and fluoroscopy</li> </ul>		
2	Heart Valve Mechanical Bi- Leaflet Mitral 23mm/25mm/27mm/29mm/31mm/33mm/All Size	<ul style="list-style-type: none"> <li>• Pyrolytic carbon Mechanical Heart Valve for Aortic Heart position.</li> <li>• Bi-Leaflet &amp; Rotatable Disk</li> <li>• U.S FDA &amp; CE approved.</li> <li>• Radio opaque for improved visualization during X-ray and fluoroscopy</li> </ul>		
3	Aotic Tissue Heart Valve 19mm/21mm/23mm/25mm/ All Size	<ul style="list-style-type: none"> <li>• Porcine Tissue Heart Valve fir Aotic Position</li> <li>• Design Tripple Composition</li> <li>• Valves have should Anticalcification Treatment</li> <li>• U.S. FDA &amp; CE approved.</li> <li>• Sewing Ring should have made of polyester with makers for helping in suturing</li> <li>• Fixation method should be low pressure Glutardehyte Fixation</li> <li>• Flexitifit stent to reduce leaflet stress</li> <li>• The Flexitifit system reduces the rinse time should be less that 50Sec.</li> </ul>		

Sr. No	Name Of Item	Specification	Name Of Company/Brand/Make	Rate Per Pcs. Inclusive All Taxes With GST
4	Mitral Tissue Heart Valve 25mm/27mm/29mm/31mm/33mm	<ul style="list-style-type: none"> <li>• Porcine Tissue Heart Valve for Mirtal Position</li> <li>• Design Tripple Composition</li> <li>• Valve have patented Anticalcification Treatment</li> <li>• Lowest Profile- 9mm</li> <li>• U.S. FDA approved.</li> <li>• Mitral Retching System facilities insertion into the Mitral annulus and avoids suture looping.</li> </ul>		
5	By-leaflet Aotic Valve with comduit 19mm,21mm/23mm/25mm/27mm/29mm/31mm	<ul style="list-style-type: none"> <li>• Should have double Velour Woven graft</li> <li>• Should be collagen Impregnated to control femostatic and reduce hemorrhagic complications</li> <li>• Should have mechanical Heart Valve with low pressure gradients with pivot guared design</li> <li>• Should have minimum tapper conduct to facilities storage coronary anastonosi</li> <li>• Should not have any Plents to allows easier position and attachement to the coronary artary</li> <li>• Should have both CE ad FDA Approval.</li> </ul>		
6	Annuoplasty Tricuspid Ring			
7	Annuoplasty Mitral Ring			
8	Retrograde Cardioplegia Cannula			
9	Mechanical Heart Valve Prosthesis Single Leaflet-Aotic and Mitral	<ul style="list-style-type: none"> <li>• Tilting Disc Valve</li> <li>• Disk Made Of High Density High Molecular Weight Polythylene</li> <li>• Low Profile</li> <li>• Roatatable in clockwiseand anticlockwise direction</li> <li>• Mounted on a disposable holder</li> <li>• MRI Compatble</li> <li>• Should be in use for more than 20 years</li> <li>• Certified by Inian FDA/DCGI/CE</li> </ul>		
10	Swan Ganz Catheter with sheath-CCO ( Cable supply free for Edward Moniter4- Lumen)			

  
 Dean,  
 Government Medical College &  
 Super Speciality Hospital,  
 Nagpur.