

महाराष्ट्र शासन

शासकीय वैद्यकीय महाविद्यालय व अतिविशेषोपचाररुग्णालय, नागपूर.

Govt. Medical College & Superspeciality Hospital, Nagpur.

Phone No.0712-2701541

Fax No.0712-2746682

E-mail-gmcssh95@gmail.com

Ref.No./GMC/SSH/PS/MJPJAY/CVTS(Heart Valve & Catheter)/QT 2172/2022

Date: - 12/08/2022

UPC

QUOTATION

To,

M/s

.....

.....

Sub :- Supply of Instrument/Equipments/Materials of the Govt. Medical College & Super Speciality Hospital, Nagpur. **Quotation for Surgical Items for CVTS Deptt. (List attached)**

Ref :- 1) GMC/SSH/PS/MJPJAY/(Heart Valve & Catheter)QT/2295-96/2021
Dated :-08/06/2021.(Validity From Date :- 01-08-2021 To 31-07-2022)
2) GMC/SSH/PS/MJPJAY/CVTS(Heart Valve & Catheter)QT/6265/2021
Dated :-08/06/2021 (Validity From Date :- 01-10-2021 To 30-09-2022)

Dear Sir,

The undersigned is invited sealed quotations for the store enlisted below/As per enclosed statement, for the use of Govt. Medical College & Super Speciality Hospital, Nagpur on the following terms & conditions.

1. The prices quoted should be for free delivery of the ordered stores to Govt. Medical College & Super Speciality Hospital, Nagpur premises F.O.R. Destination.
2. The prices quoted should be inclusive of taxes of duties including GST. If any payable like customs excise, CST, ST and G.T. the breakup of the taxes should also be shown separately wherever necessary. The Sales Tax Registration number& GST number should be quoted in your letters. Exemption of taxes, if on A.F. Forms also be separately stated.
3. The serial number of the items should not be changed while quotation rates. You may drop the item if not interested.
4. Rates should be quoted strictly for the item specified in the list & for standard quantity of goods. In case of alternate Officer, the detailed specification, name of manufacturer or make etc. must be stated. Specification other than specified in the schedule may be liable for rejection even through lowest.
5. The quotation submitted will be valid for a period of one year only from the date of acceptance.
6. Delivery period be stated specifically like ready Stock/Two weeks/Four weeks etc. & should be from supply of stores if order should be made within the stipulated period mentioned there in. Failures will be liable for delisting from further offers.

7. The quotation received after the due date will not be accepted. The quotations must be submitted in a sealed cover, unsealed quotations will not be accepted. The last date of submitting the quotation is ~~23/08/22~~ by 4:00 P.M.

8. Supply of ordered stores should be made in one installment unless otherwise ordered piecemeal supply will not be accepted. Payment will be made within 4 to 8 weeks after receive of full quantity with bills in quadruplicate & only after the satisfactory report of working etc. of the stores. Condition of advance payment through Bank. Etc. or part payment will not be accepted.

9. Quotations if asked with samples & if not accompanying the same will be liable for rejection even they are lowest samples should be sent with a tables attached quoting our ref. No. of Inquiry & Item No. etc.

10. The Govt. Medical College & Super Speciality Hospital, Nagpur do not pledge himself to accept the lowest or any quotation and reserve to right of acceptance of any quotation which suits to his requirements.


11. Incase you are not interested in quoting your rates in reply to his inquiry a line in reply is must. Failing which are liable for desisting their names for further esquires from our margin list and No further request in this matter will be entertained.

12. Literature, instructions manuals showing specifications, working etc. may also be sent with quotations.


13. Copy of Firm Registration Certificate/ST/CST/Registration No. GST No.etc. should be mentioned in the quotation while submitting the same otherwise it will treat as invalid.

14. Bidder must show Quoted surgical items to concern HOD within 3 - 5 days after stipulated period (i.e. after final submission of quotation date) for technical approval. If you failed to provide the sample within stipulated period you will be considered disqualified and will not be considered for further quotation process.

15. Technical Bid & Commercial Bid must be Submitted Separately in sealed cover, at the Dean Office GMC, Nagpur in a special quotation box of Super Speciality Hospital.



Dean,
Government Medical College &
Superspeciality Hospital, Nagpur.



CVTS DEPARTMENT QUOTATION

Ref.No./GMC/SSH/PS/MJPJAY/CVTS(Hearth Valve & Catheter)/QT/2172/2022

Date :- 12.08.22

Sr. No	Surgical Items Requirement Quotation Technical Bid	Specifition	Name Of Company	Rate Per PC Including All Taxes with GST
1	Heart Valve Mechanical Bi- Leaflet Aortic 17mm/19mm/21mm/23mm/ 25mm/All Sizes	<ul style="list-style-type: none"> Pyrolytic carbon Mechanical Heart Valve for Aortic Heart position. Bi-Leaflet U.S FDA & CE approved. Radio opaque for improved visualization during X-ray and fluoroscopy 		
2	Heart Valve Mechanical Bi- Leaflet Mitral 23mm/25mm/27mm/29mm/ 31mm/33mm/ all Sizes	<ul style="list-style-type: none"> Pyrolytic carbon Mechanical Heart Valve for both Mitral Heart position. Bi-Leaflet & Rotatable disk. U.S FDA & CE approved. Radio opaque for improved visualization during X-ray and fluoroscopy. 		
3	Aortic Tissue Heart Valve 19mm/21mm/23mm/25mm/ All Sizes	<ul style="list-style-type: none"> Porcine tissue heart valve for Aortic position. Design: Triple Composite Valves have should Anti-calcification treatment U.S. FDA approved. Sewing ring should have made of polyester with makers for helping in suturing. Fixation method should be low pressure glutardehyde fixation Flexifit stent to reduce leaflet stress The flexifit system reduces the rinse time should be less than 15 seconds. 		
4	Mitral Tissue Heart Valve 25mm/27mm/29mm/31mm/33 mm	<ul style="list-style-type: none"> Porcine tissue heart valve for Mitral position. Design: Triple Composite Valves have patented Anti-calcification treatment Lowest profile-9mm U.S. FDA approved. Mitral Retching system facilitates insertion into the mitral annulus and avoids suture looping. 		

CVTS DEPARTMENT QUOTATION

Ref.No./GMC/SSH/PS/MJPJAY/CVTS(Hearth Valve & Catheter)/QT/2172/2022

Date :- 12.08.22

5	Bi-leaflet Aortic Valve with Conduit 19mm/21mm/23mm/25mm/ 27mm/29mm/31mm	<ul style="list-style-type: none"> • Should have double velour woven Graft. • Should be collagen impregnated to control haemostatic & reduce hemorrhagic complications. • Should have Mechanical Heart Valve with low pressure gradients, with Pivot Guard design. • Should have minimum taper conduit to facilitate strong coronary anastomosis. • Should not have any pleats to allow easier position and attachment to the coronary artery. • Should have both CE & FDA approval 		
6	Annuoplasty Tricuspid Ring			
7	Annuoplasty Mitral Ring			
8	Aortic Punch All Sizes			
9	Femoral Arterial Cannula 12G			
10	Femoral Arterial Cannula 14G			
11	Femoral Arterial Cannula 15G			
12	Femoral Arterial Cannula 17G			
13	Femoral Arterial Cannula 19G			
14	Femoral Arterial Cannula 21G			
15	Femoral Venous Cannula 15G			
16	Femoral Venous Cannula 19G			
17	Femoral Venous Cannula 21G			
18	Femoral Venous Cannula 23G			
19	Femoral Venous Cannula 25G			
20	Femoral Venous Cannula 27G			
21	Arterial Cannula 18G			
22	Arterial Cannula 20G			
23	Arterial Cannula 22G			
24	Arterial Cannula 24G			
25	Left Heart Vent Catheter 10Fr			
26	Left Heart Vent Catheter 12Fr			
27	Left Heart Vent Catheter 13Fr			
28	Left Heart Vent Catheter 16Fr			
29	Left Heart Vent Catheter 17Fr			
30	Retrograde Cardioplegia Cannula			

CVTS DEPARTMENT QUOTATION

Ref.No./GMC/SSH/PS/MJPJAY/CVTS(Hearth Valve & Catheter)/QT/2172/2022

Date :- 12.08.22

Date :- 12.08.22	
31	<p>Mechanical Heart Valve Prosthesis Single Leaflet – Aortic and Mitral</p> <p>ALL SIZES</p> <ul style="list-style-type: none"> • Tilting Disc Valve • Disc made of High Density High Molecular Weight Polyethylene • Low Profile • Rotatable in clockwise and anticlockwise direction • Mounted on a disposable holder • MRI Compatible • Should be in use for more than 20 years • Certified by Indian FDA/DCGI/CE
32	<p>Swan Ganz Catheter with Sheath – CCO (cable supply free for Edward monitor 4-lumen)</p>

Dean

Government Medical College &
Superspeciality Hospital, Nagpur.

~~4/8/22~~