



**Government Medical College & Hospital, Nagpur**

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर

**Mahatma Jyotirao Phule Jan Arogya Yojana**

महात्मा ज्योतिराव फुले जन आरोग्य योजना

कार्यालयाचा पत्ता : औषधालयाचे बाजूला, शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर जेन क्र. ०७१२-२७०११००

No.GMCHN/MJPJAY/Quot/ **Cochlear Implant** / III rd Time/ **375** /2026

Date:-**25**/02/2026

To,

Subject :- **Quotation Enquiry for purchase of Cochlear Implant 3 rd Time Recall for ENT Dept through MJPJAY office.**

The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the MJPJAY, Govt. Medical College & Hospital, Nagpur under the following terms & conditions.

- 1) The prices quoted should be free delivery to MJPJAY Store, Govt. Medical College & Hospital premises **purely on Patient to Patient basis** for local dealers and F.O.R. Destination for outsiders.
- 2) Approved (L<sub>1</sub>) hardware should be made available within 24 hrs for elective cases.
- 3) Approved (L<sub>1</sub>) hardware should be available immediately for emergency cases under ETI
- 4) L<sub>2</sub>, L<sub>3</sub> or other approved vendors should supply hardware in case L<sub>1</sub> is not able to supply the hardware. (As per Government Rule)
- 5) The Price quoted should be inclusive of all taxes, duties if payable like customs, excise, GST. The GST number should be quoted in your letter while quoting the rates. Exemption of taxes, if on A.F. Forms etc. be separately stated.
- 6) The serial number of the items should not be changed while quoting rates. You may drop the item if not interested & List should be prepared in printed copy only, otherwise it may be rejected.
- 7) Rates should be quoted strictly for the item specified in the list provided and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 8) The quotation submitted will be valid for a period of **One Year** only, from the date of acceptance.
- 9) Loose packing of material will not be acceptable in any condition.
- 10) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover. Unsealed quotations, which are not properly sealed, will not be accepted. The last receiving date of the quotation is **05-03-26** up to 5.00 p.m. at MJPJAY Office Govt **Medical College & Hospital Nagpur.**
- 11) Material Should be supply in store (MJPJAY) & Your invoice and challan should have the certification that, the drug/material supplied under this challan & invoice are of required Pharmacopoeia standard and any defect found in future shall be sole responsibility of supplier.
- 12) Supply should be made from WHO GMP / FDA certified company only.
- 13) Quotations if asked with samples, if not accompanies with sample will be liable for rejection even they are lowest. Samples should be sent with a labels attached quoting our ref.no of enquiry and item no etc.

14) The Dean, Govt. Medical College & Hospital, Nagpur does not pledge himself to accept the lowest or any quotations and reserves to himself right of acceptance or rejection any of quotations which does not suit to his requirements.

15) **Very Important :- Please provide valid copy of FDA approval for Vendor as well Manufacturer.**

16) Once in a year along with the quotation please provide attested photocopies of renewed PAN No, GST/CGST/SGST, VAT, CST, Shop establishment, Drug license, Income tax clearance & etc.

17) **Very Important :- For Consumables ISO 17025,15045001,ISO 14001, GMP (GoodManufactureIngPractises) Schedule M,Quality Management System (QMS) for Medical Devices Central Drugs Standard Control Organisation (CDSCO) approved MD License.**

18) Bider should be Submit FDA's Non-Conviction Certificate.

IMP.: 1) Submit sealed quotation under Two Envelope (Two bid) system for supply of implants.  
2) Envelope No. 1 (Technical Bid : Specifications), Compliance Chart & Authorization letter, Other Certifications & Necessary Documents,

Sr No	Item Name	Manufacturer

3) Envelope No. 2 (Price Bid) submit items in format mentioned below.

Sr No	Item Name	Manufacturer	Rates per unit (Including GST)

**DEPARTMENT OF ENT, GOVT MEDICAL COLLEGE AND HOSPITAL, NAGPUR**

Technical specification for purchase of cochlear implant under Mahatma Joytirao Fule Jan Aarogya Yojana


**PRODUCT SPECIFICATION COMPLIANCE**

A				
S.No.	Product	Desired specification of TIA	Manufacture	Rate Per Unit Including GST
1)	Implant	Titanium casing of Non-Ceramic material		
2)	Intra-Cochlear Electrode Array	Multi-Channel /Multi Electrode with minimum of 12 electrode Slim Straight Electrode array or lateral wallet electrode array Minimum length of active Electrode should not be more than 25mm, Maximum diameter at the apex not more than 0.6		
3)	Speech Processor	Multi-speed coding strategy capability utilizing digitally		

	(Advanced SP with option of Body worn and BTE)	<p>programmed frequency band (12 or more)</p> <p>Should have direct streaming with Phone without any intermediate accessory/device.</p> <p>Should have control of sensitivity, volume, alarm &amp; auto features.</p> <p>Should have Mobile App application to monitor and control volume &amp; change program.</p> <p>Wide range of input processing option for clarity of sound in different environments like soft sound, Noise, Music and focus listening.</p> <p>Protection against humidity &amp; sweat.</p> <p>External speech processor of Cochlear Implant should have single/dual microphone facility</p>		
4)	Power source	Powered by rechargeable cells / rechargeable battery packs		
5)	Implant Integrity Testing ability	<p>Ability to test implant and electrode integrity and function both intraoperative and postoperatively.</p> <p>Feature for testing intra op implant functioning</p> <p>Company should provide one working equipment for automatic intra operative telemetry/compound action potential recording during entire time of the tender</p>		
6)	Speech Coding strategy	Proven and internationally accepted speech coding strategy.		
7)	MRI compatibility	MRI Compatibility- Implant should be MRI Compatibility with 1.5 T without removal of the magnet. Implant should have option of removable magnet for 3.0 T or higher MRI requirements/compatibility.		

8)	FDA Approved	Complete implant system should be US FDA Approval for implantation in paediatric and adult patients.		
9)	Feature of Implant	Implant should be compatible with current and future technology in speech processor and if company is no longer manufacturing the current processor, then it should be upgraded to higher model free of cost if within warranty period of internal device		
10)	External components	Robust and long lasting. Should be sweat/splash resistant. Easily available spare parts like magnet coil, wire and microphone sound processor etc		
11)	FM compatibility	Compatible with standard FM bases group Therapy Systems / FM		
12)	Telephone / Telecoil compatibility	Telephone compatibility with inbuilt snap/ on Telephone Adapter		
13)	Product Support	For the life of the product including implanted and external components, software and hardware.  Free upgrade of software/ hardware support for product in case existing implant system is discontinued by manufacturer.		
14)	Infrastructure support	All necessary hardware and software required for preoperative intra-operative and post operative testing and programming of the Implant should be provided free of cost to existing implant centers and any centers established in the future.		
15)	Service network	Provision of spares/loaners for equipment defects/ malfunctions Within 48 hours of request for such support  Should have wide network of		

		service and repair centers all over India to provide after sales service like repair/replacement within India of the warrantable components which includes wire, microphone, sound processor, any other accessories like magnet coil etc. to the patients residing in various parts of India, directly by the company.		
16)	Comprehensive replacement/ repair warranty is to be provided as follows.	Cochlear Implant: 10 years from the date of surgery  Processor: 3 years from date of Switch On  Accessories: 1 year for battery module, cable, coil from the date of switch on.		
17)	Training	Onsite training of medical and para medical person & continuing education program to be provided by the manufacturer as and when required		
18)	Implant	Company should have provision of return in case of non-utilization of implant (personal or surgical cause)		
19)	Kit Set.	Accessories & spares – Should be quoted with the following minimum set accessories. Coil & cables – 2 Rechargeable batteries – 02 Nos. and charger Electronic dehumidifier Interchangeable magnets of varying strength Disposable batteries – 01		



Dean  
Govt. Medical College & Hospital,  
Nagpur