

महाराष्ट्र शासन
शासकीय वैद्यकीय महाविद्यालय व अतिविशेषोपचार रुग्णालय, नागपूर.
Govt. Medical College & Super Speciality Hospital, Nagpur.

Phone No. 0712-2701541

Fax No. 0712-2746682

E-mail-gmcssh95@gmail.com

Ref.No./GMC & SSH/PS/CSSD /QT/

/2026

Date :- 13/8 /2026

7463

UPC

QUOTATION

To,

M/s

.....

Sub :- Supply of Instrument/Equipment/Materials of the Govt. Medical College & Super Speciality Hospital, Nagpur (**Quotation for Surgical Items required at CSSD Department.**)

Dear Sir,

The undersigned is invited sealed quotations for the store enlisted below/As per enclosed statement, for the use of Govt. Medical College & Super Speciality Hospital, Nagpur on the following terms & conditions.

1. The prices quoted should be for free delivery of the ordered stores to Govt. Medical College & Super Speciality Hospital, Nagpur premises F.O.R. Destination.
2. The prices quoted should be inclusive of taxes of duties including GST. If any payable like customs excise, CST, ST and G.T. the breakup of the taxes should also be shown separately wherever necessary. The Sales Tax Registration number & GST number should be quoted in your letters. Exemption of taxes, if on A.F. Forms also be separately stated.
3. The serial number of the items should not be changed while quotation rates. You may drop the item if not interested.
4. Rates should be quoted strictly for the item specified in the list & for standard quantity of goods. In case of alternate Officer, the detailed specification, name of manufacturer or make etc. must be stated. Specification other than specified in the schedule may be liable for rejection even through lowest.
5. The quotation submitted will be valid for a period of one year only from the date of acceptance.
6. Delivery period be stated specifically like ready Stock/Two weeks/Four weeks etc. & should be from supply of stores if order should be made within the stipulated period mentioned there in. Failures will be liable for delisting from further offers.
7. The quotation received after the due date will not be accepted. The quotations must be submitted in a sealed cover, unsealed quotations will not be accepted. **The last date of receiving the quotation is 19/08/2026 by 4:00 P.M.**
8. Supply of ordered stores should be made in one installment unless otherwise ordered piecemeal supply will not be accepted. Payment will be made within 4 to 8 weeks after receive of full quantity

with bills in quadruplicate & only after the satisfactory report of working etc. of the stores. Condition of advance payment through Bank. Etc. or part payment will not be accepted.

9. Quotations if asked with samples & if not accompanying the same will be liable for rejection even they are lowest samples should be sent with a tables attached quoting our ref. No. of Inquiry & Item No. etc.

10. The Govt. Medical College & Super Speciality Hospital, Nagpur do not pledge himself to accept the lowest or any quotation and reserve to right of acceptance of any quotation which suits to his requirements.

11. In case you are not interested in quoting your rates in reply to his inquiry a line in reply is must. Failing which are liable for desisting their names for further esquires from our margin list and No further request in this matter will be entertained.

12. Literature, instructions manuals showing specifications, working etc. may also be sent with quotations.

13. Copy of Firm Registration Certificate/ST/CST/Registration No. GST No etc. should be mentioned in the quotation while submitting the same otherwise it will treat as invalid.

14. Bidder must show Quoted surgical items to concern HOD within 3 - 5 days after stipulated period (i.e. after final submission of quotation date) for technical approval. If you failed to provide the sample within stipulated period you will be considered disqualified and will not be considered for further quotation process.

15. Technical Bid & Commercial Bid must be submitted with separately in sealed cover.

(Note: - The required information about the Items/materials/Consumables etc., has to be filled Separately for Technical Bid & Commercial Bid in or as per the given table format.)

Digitally signed by
RAJ NARENDRA GAJBHIYE
Date: 12-08-2026
21:59:49

(Dr. Raj Gajbhiye)
Dean,
Government Medical College &
Super Speciality Hospital,
Nagpur.

महाराष्ट्र शासन
शासकीय वैद्यकीय महाविद्यालय व अतिविशेषोपचार रुग्णालय, नागपूर.
Govt. Medical College & Super Speciality Hospital, Nagpur.

Phone No.0712-2701541 Fax No.0712-2746682 E-mail-gmcssh95@gmail.com
 Ref.No./GMC&SSH/PS/CSSD/QT/7463 /2026 Date :- 13 / 8 /2026

TECHNICAL BID FORM

CSSD Department Requirement Quotation

(It is Compulsory to provide sample for Technical Approval)

Sr.No	Name Of Item	Brand/Make/Company Name
1	Class 1 Indicators STEAM - 3 Line Double Adhesive STEAM label - color change after sterilization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 9001, ISO 13485.	
2	Class 1 Indicators ETO - 3 Line Double Adhesive ETO label - color change after sterilization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 13485	
3	Class 1 Indicators H2O2 Plasma - 3 Line Double Adhesive PLASMA Label - after sterilization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 9001, ISO 13485	
4	Monarch 1153 - 3 Line Labelling gun with Ink roll - 1 nos 1 Gun Given Free of cost If all Class 1 Indic. Provided in Order to us	
5	Class 2 Indicators STEAM- Bowie Dick test Hollow PCD 4 Meters Lumen length, 2 mm diameter, SS Device, Color Change from Yellow to Black in 6 separate block indicators. Non-Toxic, Lead Free, EN 285, EN ISO 11140 - Part 1:2009 and Part 4:2007 adherence. ISO 9001, ISO 13485.	
6	Class 2 Indicators STEAM- Bowie Dick test Hollow PCD 4 Meters Lumen length, 2 mm diameter, SS Device, Color Change from Yellow to Black in 6 separate block indicators. Non-Toxic, Lead Free, EN 285, EN ISO 11140 - Part 1:2009 and Part 4:2007 adherence. ISO 9001, ISO 13485. 1 PCD - Will be given Free in consignment	
7	Class 4 Indicator - PLASMA, Package Monitoring Indicator, Color change after sterilization 6 Minutes, 50 Degrees, 2.3 mg/l H2O2, 100% Toxic and Metal Free, ISO 11140-1, ANMAT PM 1614-1, ISO 9001, 13485, Shelf Life - 5 Years	
8	Class 5 Indicators - Biological Indicators - Steam Self Contained 1 x 10⁶ Population with Certificate of Analysis of each Batch. ISO 11138-1, ISO 11138-3, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value not lower than 1.5 min (121 Degrees C)	

9	Class 5 Indicators - Biological Indicators - ETO Self Contained 1 x 10 ⁶ Population with Certificate of Analysis of each Batch. ISO 11138-1, ISO 11138-3, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value not lower than 2.5 min (600 mg/l EO))	
10	Class 5 Indicators - Biological Indicators - PLASMA Self Contained 1 x 10 ⁶ Population with Certificate of Analysis of each Batch. ISO 11138-1, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value 2 mg/l H2O2	
11	Class 6 Indicators STEAM - Emulating Indicator " Not Sterile" Mentioned with clear color change from Yellow to black only, With Ref Color Change, With Self-Adhesive Sticking, Double Laminated, Non-Toxic, Lead-Free, Metal Free. ISO 11140-1 Type 6 and Lot Number should be mentioned on Each Indicator for Lot Tracking. ISO 9001, ISO 13485	
12	Class 5 Integrator ETO- ETO indicator 60 x 25 mm, with color change from Red and Brown to Yellow green only, 100% Toxic and Metal Free Substances, ISO 11140-1, Type 5, ISO 9001, ISO 13485.	
13	Class 5 Indicator ETO- Batch Monitoring System (BMS) - Lot Control Test with Helix PCD of 1.5 Meters Length, 2 mm lumen dia, Teflon coated. 4 Block Indicator Color Change from light Blue to Green. 100% Toxic and Metal Free Substances, EN 285, EN 867-5, ISO 11140-1, ISO 11140-4, ISO 11140-5 Standards, ISO 13485 Certified	
14	Ultrasonic Cleaner Indicator For Monitoring Ultrasonic Washer High Challenge Level, Color elimination of Funchsia organic substance, Size 70mm x 30mm , ISO/TS 15883-5,EN ISO 13485,5 Years shelf life ,	
15	Packing Reel - 05 cm x 200 mts	
16	Packing Reel - 7.5 cm x 200 mts	
17	Packing Reel - 10 cm x 200 mts	
18	Packing Reel - 15 cm x 200 mts	
19	Packing Reel - 20 cm x 200 mts	
20	Packing Reel - 25 cm x 200 mts	
21	Packing Reel - 30 cm x 200 mts	
22	Packing Reel - 35 cm x 200 mts	
23	Packing Reel - 40cm x 200 mts	
24	Tyvek Sterilization Roll 50 mm x 70 M	
25	Tyvek Sterilization Roll 75mm x 70 M	
26	Tyvek Sterilization Roll 100 mm x 70 M	
27	Tyvek Sterilization Roll 150 mm x 70 M	
28	Tyvek Sterilization Roll 200 mm x 70 M	
29	Tyvek Sterilization Roll 250 mm x 70 M	
30	Tyvek Sterilization Roll 300 mm x 70 M	
31	Tyvek Sterilization Roll 350 mm x 70 M	
32	Kinguard - 61 cm x 61 cm	
33	Kinguard - 76 cm x 76 cm	
34	Kinguard - 101 cm x 101 cm	
35	Kinguard - 121 cm x 121 cm	

36	<u>Multi-Enzyme Solution</u> 12% v/v Protease, Lipase, Amalyase with Non-Ionic Surfactant, PH Neutral, Corrosion inhibitors, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485	
37	<u>Rust & Stain Remover</u> Inorganic Acid (> 30% Phosphoric Acid), Non Ionic Surfactant (<5%), Non-flammable, No harmful VOC's, GMP Certified, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485	
38	<u>High-Level Disinfectant</u> Aldehyde Free - 3% w/v [poly(hex methylene biguanide)hydrochloride] (PHMB) , 10% w/v Dodecyl dimethyl ammonium chloride (DDAC) with Sporicidal and Mycobactericidal activity, Colorless, odorless, known action against nosocomial infections, biofilms, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485	
39	<u>Neutralizer</u> for Machine Cleaning the thermostable & thermolabile instrument & laboratory glass ware	
40	Alkaline resistant cleaner for metal, glass, plastic, rubber Eliminate area & dirt (blood, protein, oil, grease, resin, wax, thoroughly.)	
41	Abro Tape	
42	Autoclave Tape	
43	Ink Caset(ETO)	
44	Thermal Printer Paper for Steam Sterlizer (110MM)	
45	<u>Cleaning Brushes</u> Instrument Cleaning Brushes - Set of 2 pcs (Double Sided Brushes with 1 nos Lumen Device and 1 nos Regular brush) for cleaning Suction/ scopes/ lumen devices	
46	<u>SIMPLE PPE Kit – Disposable</u> Personal Protection Kit for WASHING Area with Gown, Gloves, Waste bag, Goggle, Head Cap, Face Mask, Shoe Cover	
47	<u>Fluid shield Facemask with Visor</u> PFE > 99%, BFE > 99%, ASTM 160mm Hg, Ear loop, Foam Band, Orange Color	
48	<u>Medical Safety Clogs</u> Safety Shoes - Covered with Aeration Holes, Anti-Static, Anti-Slip, Chemical Resistant, Good Grip, Blue Colored, Non-Flammable, Washable	
49	<u>Safety Gloves for Wet Steam Handling for Autoclave</u> Upto 300 Degrees safety, Silicon Material, Flexible, water-proof, Washable, Imported with all Certificates of Analysis	

50	Special CSSD Marker Pen Non-Toxic Ink, Permanent Marker for Autoclave/ ETO/ Plasma. BLACK Color	
51	Temperature & Humidity Meter Sterile Area Monitor for Temperature, Humidity and time	
52	Stain and Rust Remover Pad Scotch brite type safe pad for removal of rust, stain, blood from soiled instruments, Nondamaging to instruments, Brown Color, Approx Size - 150 x 100 mm	
53	Multimat Disinfectant Multimat - 60 x 115 cm - Blue with Antimicrobial coating - For Critical area, (30 Layers in 1 Mat set)	
54	Door Seal (Steam)	
55	Bacterial Filter (Steam)	
56	Plasma cassette	
57	Printing Paper Roll (Plasma) 75mm	
58	Printing Paper Roll (ETO) 58mm x 15 mtr)	
59	Cartridge+Humidi Chip+ Dosimeter	
60	Micron Filter (Ro)	
61	Membran (RO)	
62	Antisacilent Liquid	

Digitally signed by
RAJ NARENDRA GAJBHIYE
Date: 12-08-2026
21:58:55

(Dr. Raj Gajbhiye)
Dean,
Government Medical College &
Super Speciality Hospital,
Nagpur.

महाराष्ट्र शासन
शासकीय वैद्यकीय महाविद्यालय व अतिविशेषोपचार रुग्णालय, नागपूर.
Govt. Medical College & Super Speciality Hospital, Nagpur.

Phone No.0712-2701541

Fax No.0712-2746682

E-mail-gmcssh97@gmail.com

Ref.No./GMC&SSH/PS/CSSD/QT/ 7483 /2026

Date :- 13/08/2026

PRICE BID FORM

CSSD Department Requirement Quotation

(It is Compulsory to provide sample for Technical Approval)

Sr.No	Name Of Item	Brand/Make/Company Name
1	Class 1 Indicators STEAM - 3 Line Double Adhesive STEAM label - color change after sterilization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 9001, ISO 13485.	
2	Class 1 Indicators ETO - 3 Line Double Adhesive ETO label - color change after sterilization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 13485	
3	Class 1 Indicators H2O2 Plasma - 3 Line Double Adhesive PLASMA Label - after sterilization Non-Toxic, Lead Free, ISO 11140-1, Type-1 mentioned on Label. Roll of 750 Labels, each Roll with Lot Number, Expiry mentioned. ISO 9001, ISO 13485	
4	Monarch 1153 - 3 Line Labelling gun with Ink roll - 1 nos 1 Gun Given Free of cost If all Class 1 Indic. Provided in Order to us	
5	Class 2 Indicators STEAM- Bowie Dick test Hollow PCD 4 Meters Lumen length, 2 mm diameter, SS Device, Color Change from Yellow to Black in 6 separate block indicators. Non-Toxic, Lead Free, EN 285, EN ISO 11140 - Part 1:2009 and Part 4:2007 adherence. ISO 9001, ISO 13485.	
6	Class 2 Indicators STEAM- Bowie Dick test Hollow PCD 4 Meters Lumen length, 2 mm diameter, SS Device, Color Change from Yellow to Black in 6 separate block indicators. Non-Toxic, Lead Free, EN 285, EN ISO 11140 - Part 1:2009 and Part 4:2007 adherence. ISO 9001, ISO 13485. 1 PCD - Will be given Free in consignment	
7	Class 4 Indicator - PLASMA, Package Monitoring Indicator, Color change after sterilization 6 Minutes, 50 Degrees, 2.3 mg/l H2O2, 100% Toxic and Metal Free, ISO 11140-1, ANMAT PM 1614-1, ISO 9001, 13485, Shelf Life - 5 Years	
8	Class 5 Indicators - Biological Indicators - Steam Self Contained 1 x 10⁶ Population with Certificate of Analysis of each Batch. ISO 11138-1, ISO 11138-3, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value not lower than 1.5 min (121 Degrees C)	

9	Class 5 Indicators - Biological Indicators - ETO Self Contained 1 x 10 ⁶ Population with Certificate of Analysis of each Batch. ISO 11138-1, ISO 11138-3, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value not lower than 2.5 min (600 mg/l EO)	
10	Class 5 Indicators - Biological Indicators - PLASMA Self Contained 1 x 10 ⁶ Population with Certificate of Analysis of each Batch. ISO 11138-1, IRAM 37102 (Part 1 & 3), ISO 13485, ANMAT PM-1614-1 and USP Compliant. D Value 2 mg/l H2O2	
11	Class 6 Indicators STEAM - Emulating Indicator " Not Sterile" Mentioned with clear color change from Yellow to black only, With Ref Color Change, With Self-Adhesive Sticking, Double Laminated, Non-Toxic, Lead-Free, Metal Free. ISO 11140-1 Type 6 and Lot Number should be mentioned on Each Indicator for Lot Tracking. ISO 9001, ISO 13485	
12	Class 5 Integrator ETO - ETO indicator 60 x 25 mm, with color change from Red and Brown to Yellow green only, 100% Toxic and Metal Free Substances, ISO 11140-1, Type 5, ISO 9001, ISO 13485.	
13	Class 5 Indicator ETO- Batch Monitoring System (BMS) - Lot Control Test with Helix PCD of 1.5 Meters Length, 2 mm lumen dia, Teflon coated. 4 Block Indicator Color Change from light Blue to Green. 100% Toxic and Metal Free Substances, EN 285, EN 867-5, ISO 11140-1, ISO 11140-4, ISO 11140-5 Standards, ISO 13485 Certified	
14	Ultrasonic Cleaner Indicator For Monitoring Ultrasonic Washer High Challenge Level, Color elimination of Funchsia organic substance, Size 70mm x 30mm , ISO/TS 15883-5,EN ISO 13485,5 Years shelf life ,	
15	Packing Reel - 05 cm x 200 mts	
16	Packing Reel - 7.5 cm x 200 mts	
17	Packing Reel - 10 cm x 200 mts	
18	Packing Reel - 15 cm x 200 mts	
19	Packing Reel - 20 cm x 200 mts	
20	Packing Reel - 25 cm x 200 mts	
21	Packing Reel - 30 cm x 200 mts	
22	Packing Reel - 35 cm x 200 mts	
23	Packing Reel - 40cm x 200 mts	
24	Tyvek Sterilization Roll 50 mm x 70 M	
25	Tyvek Sterilization Roll 75mm x 70 M	
26	Tyvek Sterilization Roll 100 mm x 70 M	
27	Tyvek Sterilization Roll 150 mm x 70 M	
28	Tyvek Sterilization Roll 200 mm x 70 M	
29	Tyvek Sterilization Roll 250 mm x 70 M	
30	Tyvek Sterilization Roll 300 mm x 70 M	
31	Tyvek Sterilization Roll 350 mm x 70 M	
32	Kinguard - 61 cm x 61 cm	
33	Kinguard - 76 cm x 76 cm	
34	Kinguard - 101 cm x 101 cm	
35	Kinguard - 121 cm x 121 cm	

36	<u>Multi-Enzyme Solution</u> 12% v/v Protease, Lipase, Amylase with Non-Ionic Surfactant, PH Neutral, Corrosion inhibitors, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485	
37	<u>Rust & Stain Remover</u> Inorganic Acid (> 30% Phosphoric Acid), Non Ionic Surfactant (<5%), Non-flammable, No harmful VOC's, GMP Certified, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485	
38	<u>High-Level Disinfectant</u> Aldehyde Free - 3% w/v [poly(hex methylene biguanide)hydrochloride] (PHMB) , 10% w/v Dodecyl dimethyl ammonium chloride (DDAC) with Sporicidal and Mycobactericidal activity, Colorless, odorless, known action against nosocomial infections, biofilms, With Validation Reports, MSDS, CoA, Dilution Charts, Training on Usage. CE (93/42/EEC Class II b or 98/8/EC), 13485	
39	<u>Neutralizer</u> for Machine Cleaning the thermostable & thermolabile instrument & laboratory glass ware	
40	Alkaline resistant cleaner for metal, glass, plastic, rubber Eliminate area & dirt (blood, protein, oil, grease, resin, wax, thoroughly.)	
41	Abro Tape	
42	Autoclave Tape	
43	Ink Caset(ETO)	
44	Thermal Printer Paper for Steam Sterilizer (110MM)	
45	<u>Cleaning Brushes</u> Instrument Cleaning Brushes - Set of 2 pcs (Double Sided Brushes with 1 nos Lumen Device and 1 nos Regular brush) for cleaning Suction/ scopes/ lumen devices	
46	<u>SIMPLE PPE Kit – Disposable</u> Personal Protection Kit for WASHING Area with Gown, Gloves, Waste bag, Goggle, Head Cap, Face Mask, Shoe Cover	
47	<u>Fluid shield Facemask with Visor</u> PFE > 99%, BFE > 99%, ASTM 160mm Hg, Ear loop, Foam Band, Orange Color	
48	<u>Medical Safety Clogs</u> Safety Shoes - Covered with Aeration Holes, Anti-Static, Anti-Slip, Chemical Resistant, Good Grip, Blue Colored, Non-Flammable, Washable	
49	<u>Safety Gloves for Wet Steam Handling for Autoclave</u> Upto 300 Degrees safety, Silicon Material, Flexible, water-proof, Washable, Imported with all Certificates of Analysis	

50	Special CSSD Marker Pen Non-Toxic Ink, Permanent Marker for Autoclave/ ETO/ Plasma. BLACK Color	
51	Temperature & Humidity Meter Sterile Area Monitor for Temperature, Humidity and time	
52	Stain and Rust Remover Pad Scotch brite type safe pad for removal of rust, stain, blood from soiled instruments, Nondamaging to instruments, Brown Color, Approx Size - 150 x 100 mm	
53	Multimat Disinfectant Multimat - 60 x 115 cm - Blue with Antimicrobial coating - For Critical area, (30 Layers in 1 Mat set)	
54	Door Seal (Steam)	
55	Bacterial Filter (Steam)	
56	Plasma cassette	
57	Printing Paper Roll (Plasma) 75mm	
58	Printing Paper Roll (ETO) 58mm x 15 mtr)	
59	Cartridge+Humidi Chip+ Dosimeter	
60	Micron Filter (Ro)	
61	Membran (RO)	
62	Antisacilent Liquid	

Digitally signed by
RAJ NARENDRA GAJBHIYE
Date: 12-08-2026
21:58:22

(Dr. Raj Gajbhiye)
Dean,
Government Medical College &
Super Speciality Hospital,
Nagpur.