

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर

GOVERNMENT MEDICAL COLLEGE & HOSPITAL, NAGPUR

FAX NO. 0712-274448

Tele. No. Office:-2743588

No.Hosp/PS/Qt/Enq./ 292-97 /2022

Dt. 12.1.2023

Quotation

UPC

To,

M/s.....

.....

.....

.....

.....

Subject :- Submission of Quotation for the Supply of Instrument/Articles/Materials to the Govt. Medical College & Hospital, Nagpur.


The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the Govt. Medical College & Hospital, Nagpur on the following terms & conditions.

- 1) The prices quoted should be free delivery to College & Hospital premises for local dealers and F.O.R. Nagpur for outsiders.
- 2) The Price quoted by inclusive of all taxes, duties if payable like customs, excise, CST, ST, and GST. The break-up of the taxes should also be shown separately where necessary. The sales tax & registration number should be quoted in your letter. Exemption of taxes, if on A.F. Forms etc.be separately stated.
- 3) The serial number of the items should not be changed while quoting rates. You may drop the item if not interested.
- 4) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 5) The quotation submitted will be valid for a period of One (01) year only from the date of acceptance.
- 6) Delivery period should be stated specifically the like ready stock or two weeks or four weeks etc and should be firm and supply of stores if ordered should be made available within the stipulated period. Failures are liable for delist from further enquiries.
- 7) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover. Unsealed quotations which is not properly sealed will not be accepted. The last or receiving date of the quotation is Dt. 20.1.2023 up to 3.00 p.m.
- 8) Supply of stores should be made in one installment unless otherwise ordered place meal supply will not be accepted. Condition of advance payment through Bank, part payment etc. will not be accepted.
- 9) The price quoted is inclusive of all taxes, duties if payable like Customs/ Excise/CST/ST the break up of the taxes should also be shown separately where necessary. The sales tax and registration number should be quoted in your letter. Exemption of taxes on 'AF'/ Octroi exemption form etc be separately stated.

(Contd. On ...2..)

...2...

- 10) Quotations if asked with samples if not accompanies with sample will be liable for rejection even they are lowest Samples should be sent with a labels attached quoting our ref.no of enquiry and item no etc.
- 11) The Dean, Govt. Medical College & Hospital, Nagpur does not pledge himself to accept the lowest or any quotations and reserves to himself right of acceptance or rejection any of quotations which does not suit to his requirements.
- 12) Very Important :- In case you are not interested in quoting your rates; in reply to this enquiry a line in reply is must; failing are liable for delisting their names for further enquiries from our list and no further request in this matter will be entertained.
- 13) Literature, Instruction manual, working leaflet etc may also be sent with quotation.
- 14) Authorization certificate, Brochure, Catalogue, User List may also submit along with quotation is mandatory.
- 15) Technical Specifications - Compliance Chart with variation/deviation.
- 16) Bidder should apply in 2 Bid – TECHNICAL BID & COMMERCIAL BID also write clearly on envelop as TECHNICAL BID and COMMERCIAL BID in BOLD letter. Bid submitted separately in sealed cover with refer to our quotation No. in TOP of Envelop otherwise their quotation not to be accepted.
- a) First Bid (Technical Bid) should including Registration/Authorization Certificate. Sale Tax, IT Return Certificate, PAN card, GST no compliance chart as per specification. Brochure of the quotes material as per specification write clearly an envelope as TECHNICAL BID.
- b) Second Bid (Commercial Bid) Should include items rate including all taxes only.
- 17) While submitting the bills of good supplied, supplier must calculate unit price and final unit price should reflect the same rate as per order given by the Institute, GST & other taxes should not be calculated on total prices.
- 18) After opening the technical BID qualified Bidder will have to give demonstration of equipment as well as show their quoted items to HOD for technical approval as a Technical Committee otherwise Bidder names not to be included in commercial bid.



DEAN

Govt. Medical College & Hospital, Nagpur

Details as below: (As per specification attached)

Sr. No.	Name of the Items	Dept.	Price (Including GST)
1	High End Cautery	ENT	

Technical specification
High End Cautery
Cautery machine with vessel sealing system

1) Unit should have microprocessor controlled tissue feedback technology with LCD touchscreen
2) It should adjust power level automatically depending on tissue type.
3) Unit should have error code function for fault conditions.
4) It should accept dual area and single area patient return electrode.
5) It should have RANDOMIZED spray coagulation mode.
6) It should have at least TEN USER SETTABLE programs for different surgical procedures.
7) Unit should be useful for underwater procedures.
8) It should have Alarm facility after completion of Bipolar coagulation.
9) Unit should have touch key pad for power and mode selection.
10) It should have digital display which should indicate true power for selected mode.
11) Unit should operate from 170V to 280V without using external stabilizer.
12) It should have separate and isolated sockets for Monopolar and Bipolar.
13) It should meet European CE or USFDA safety and design standard.
14) Company should be ISO certified.
15) It should have THREE different modes for Cutting:
PURE CUT: 300W at 300 Ohms
BLEND CUT 1: 100W at 300 Ohms, 250-300
BLEND CUT 2: 200W at 300 Ohms
It should have SEVEN different modes for Coagulation:
SOFT : 120 -300 wat-500 Ohms
FULGURATE: 120 W - 500 Ohms
Desiccate High spray
And hydro coagulation: 120 W at 500 Ohms
It should have TWO different modes for Bipolar Coagulation
MICRO : 70 - 120watt at 100 Ohms,
MACRO : 70 - 120watt at 100 Ohms
16) It should have Endo 1, Endo2 & hydro cut of 400 Watt each
17) It should be supplied with following accessories:
a) Patient return electrode-----1 No.
b) Cable for return electrode---1 No.
c) Hand switching pencil-----1 no.
d) Foot switching pencil-----1 no.
e) Bipolar forceps-----1 No.

Technical specification
High End Cautery
Cautery machine with vessel sealing system

f)	Cable for bipolar forceps----1 No.
g)	Monopolar Foot switch-----1 no.
h)	Bipolar foot switch-----1 no.
i)	Universal adaptor----- 1 No.
j)	Reusable vessel sealing clamp c_ ratchet cconnecting cable
NOTE: All accessories should be reusable and autoclavable.	
18)	It should have separate socket for vessel seal
19)	It should give alarm for instrument failure for Open and close of instrument.
20)	Vessel sealer instrument should be reusable and repairable.
21)	Vessel sealer should have up to 35 cm long lapro insert and up to 18cm short insert for open surgeries with Handle with ratchet
22)	Seal Mode: Level 1 to Level 3, 10 Ohms to 200 Ohms,
23)	Auto Stops the energy delivery with audio indication at completion of sealing.

