

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर

GOVERNMENT MEDICAL COLLEGE & HOSPITAL, NAGPUR

Tele. No. Office:-2743588

FAX NO. 0712-274448

No.Hosp/PS/QT/Enq/

8030 /2022

Dt. 22/4 /2022

Quotation

UPC

To,
M/s.....
.....
.....

Subject: - Submission of Quotation for the Supply of Instrument/Articles/Materials to the Govt. Medical College & Hospital, Nagpur.

The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the Govt. Medical College & Hospital, Nagpur on the following terms & conditions.

- 1) The prices quoted should be for delivery to College & Hospital premises for local dealers and F.O.R: Nagpur for outsiders.
- 2) The Price quoted by inclusive of all taxes, duties if payable like customs, excise, CST, ST, and GST. The break-up of the taxes should also be shown separately wherever necessary. The sales tax & registration number should be quoted in your letter. Exemption of taxes, if on A.F. Forms etc.be separately stated.
- 3) **The serial number of the items should not be changed while quoting rates.** You may drop the item if not interested.
- 4) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 5) The quotation submitted will be valid for a period of One (01) year only from the date of Acceptance.
- 6) Delivery period should be stated specifically , like ready stock or two weeks or four weeks etc and should be made available to this store within the stipulated period. Failures are liable for delist from further enquiries.
- 7) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover and in **separate envelopes** labelled clearly as **"BID 1- Technical" and "BID 2 – Commercial"** Unsealed quotations which are not properly sealed will not be accepted. The last or receiving date of the quotation is Dt. 29/04/2022 upto 3.00 p.m.
- 8) Supply of stores should be made in one instalment unless ordered otherwise. Piecemeal supply will not be accepted. Condition of advance payment through Bank, partpayment etc. will not be accepted.
- 9) The price quoted is inclusive of all taxes, duties if payable like Customs/ Excise/CST/ST the breakup of the taxes should also be shown separately where necessary. The sales tax and registration number should be quoted in your letter. Exemption of taxes on 'AF/ Octroi exemption form etc.is separately stated.

(Contd. On ...2..)

...2...

10) Quotations if asked with samples should be accompanied if not accompanied with sample it will be liable for rejection even they are lowest. Samples should be sent with a labels attached quoting our ref.no of enquiry and item no etc.

11) The Dean, Govt. Medical College & Hospital, Nagpur does not pledge himself to accept the lowest or any quotations and reserves the right of acceptance or rejection of any quotations which are not as per the specification or requirement .

12) Supplier must quote the rates of items only those which he/she will display/demonstrate and supply. If the good supplied are different from the sample displayed or checked, order will be cancelled & no payment will be done.

13) Literature, Instruction manual, working leaflet etc may also be sent with quotation.

14) Registration/Authorization Certificate, Sale tax, IT Return Certificate, PAN card, GST no Compliance chart ,brochure ,Catalogue, User List submission (wherever applicable) along With quotation is mandatory.

15) Technical Specifications - Compliance Chart with variation/deviation

16) Bidder should apply in 2 Bid – TECHNICAL BID & COMMERCIAL BID also write clearly on envelop as TECHNICAL BID and COMMERCIAL BID in BOLD Letter. Bid submitted separately in sealed cover with reference to our Quotation No. on Top of Envelope otherwise their quotation not to be accepted.

(a) First Bid (Technical Bid). Should including Registration/Authorization Certificate, Sale tax, IT Return Certificate, PAN card, GST no Compliance chart as per specification, Brochure of the quotes Material as per specification write clearly an envelope as TECHNICAL BID.

(b) Second Bid (commercial Bid). Should include items rate including all Taxes only.

17) While submitting the bills of good supplied, supplier must calculate unit price and final unit price should reflect the same rate as per order given by the Institute. GST & other taxes should not be calculated on total prices.

18) After Opening the technical BD qualified Bidder will have to give demonstration of equipment as well as show their quoted items to HOD for technical approval as a Technical Committee otherwise Bidder names not to be included in commercial bid.



DEAN
Govt. Medical College & Hospital,
Nagpur

Details as Below:

Sr.No	Name of Item with specification	Qty	Rate Incl.GST	Department
1	Neuromuscular Monitor/TDF Monitor	1Nos		Anesthesia
2	Flexible Intubation Video Endoscope (Pediatric) Single Use compatible with existing C MAC monitor in department (C MAC Monitor (8403ZX) & C-HUBII920290320) (Quantity -03) with one E box for Karl Storz flexible single use video endoscope connector compatible with C MAC Monitor(Quantity -01)	One Set		

Note: Rate are Including GST and all taxes.

Specifications for Flexible Intubation Video Endoscope (Ped)- Single Use – compatible to be use with existing CMAC monitor(Portable)

1- Flexible Intubation Video Endoscope – Single Use – Pediatric – Qty 3


- It should be light weight, high resolution & portable flexible scope for single use
- Integrated camera chip and LED light illumination
- ET TUBE HOLDER should be a part of standard accessory of flexible scope and should be from same manufacturer.
- Each scope should be in a separate sterile packing and ready for use
- Should be supplied with Connector for Single Use flexible scopes to be used with existing portable Monitor in the department CMAC Monitor (8403ZX) & C- HUBII920290320) or separate monitor/ camera processor to be supplied for Image visualization.
- Flexible intubation video endoscope should be compatible with existing monitor in department CMAC Monitor (8403ZX) & C- HUBII920290320)
- Scope should be suitable for following applications-
 - Endotracheal Intubation
 - Bronchoscopy
 - Foreign body removal
 - Bronchial Lavage
 - Inspection of the Airways
 - Dilatation Tracheotomy
- Technical Details of Flexible Video Endoscope-
 - Tip deflection UP/DOWN: 180° or more/180° or more
 - Angle of view 90-100 degrees,
 - Working Length: minimum 65 cm or more
 - Total length: at least 90-100 cm,
 - Working Channel diameter: 1.2 mm or more,
 - Distal Tip Outer Diameter: 3.4mm-3.6mm


All above items should be USFDA and European CE certified only.


All the above items should be from same parent company .

Above supplied material should be compatible with existing portable monitor in department.

Prior Demonstration of equipment is must


Professor & Head
Dept. of Anaesthesia
Govt. Medical College Nagpur


Assist. / Assoc. / Prof.
Dept. of Anaesthesiology
Govt. Medical College &
Hospital, Nagpur


Assist. / Assoc. / Prof.
Dept. of Anaesthesiology
Govt. Medical College &
Hospital, Nagpur

NEUROMUSCULAR MONITOR SPECIFICATION

- It should have various stimulation modes like TOF, PTC, TET, DBS, SINGLE TWITCH
- It should have 3 D acceleration measurement battery should give continuous at 1500 hrs of operation, it should have standard AA battery 9a; alkaline or NiCd or NiMH)
- It should have at least large 4.4 inch display, LCD type with resolution 240X320
- It should be integrated to patient monitor & operation via patient monitor
- It should have arrangement to place at operation table, IV stand or standard rail
- It should have warning messages for lower & upper TOF limit
- It should measure surface body temperature
- It should have integrated trend memory that can be downloaded to external devices, memory in capacity should have 180 hours of recording
- Stimulation it should have monophasic rectangular waveform with pulse width 2300 to 300 micro sec & constant current of 0-60 mA, load 100 ohm to 3 k ohm
- Electrical specification it should be internally powered handheld device type BF, continuous operation, IPX, operating voltage 4 to 6 volts max current 330 MA, power consumption max 2.5 Watt
- Dimensions approximately 62 X 140 X 200 mm
- Demonstration of equipment is must

Sabshi
29.3.22
Assist. / Assoc. / Prof.
Dept. of Anaesthesiology
Govt. Medical College &
Hospital, Nagpur

AN
29.3.22
Professor & Head
Dept. of Anesthesia
Govt. Medical College, Nagpur

YN
29.3.22
Assist. / Assoc. / Prof.
Dept. of Anaesthesiology
Govt. Medical College &
Hospital, Nagpur