

महाराष्ट्र शासन
शासकीय वैद्यकीय महाविद्यालय व अतिविशेषीयचारुम्हालय, नागपूर.
Govt. Medical College & Superspeciality Hospital, Nagpur.

Phone No. 0712-2791541

Fax No. 0712-2791582

E-mail: gmcssh95@gmail.com

Ref.No./GMC&SSH/PS/BLOOD BANK/QT/ 642 /2022

Date: -22/2/22

**Invitation of quotation for
Laminar Air flow at Blood Bank Department, GMC & Super
Speciality Hospital, Nagpur.**

Sealed Quotations are hereby invited by the undersigned on behalf of the Dean, GMC & Super Speciality Hospital, Nagpur for Laminar Air flow for the Institute as per terms & condition mentioned below. The filled quotation along with all the required document must reach in the Dean Office at Govt. Medical College, Nagpur on or before Dt:- 28/02/2022, 04:00PM. The Envelope containing the quotation would please be sealed and super scribed as under

**"QUOTATION FOR LAMINAR AIR FLOW
DUE ON DT:-28/02/2022 BY 4:00 PM"**

1. Terms & Conditions:

The undersigned is invited sealed quotations for the store enlisted below/as per enclosed statement, for the use of Govt. Medical College & Super Speciality Hospital, Nagpur on the following terms & conditions.

- A) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in "Quotation Box" located in Office of Dean, GMC, Nagpur.
- B) The quotations received after deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. The offer Submitted by Fax/Email shall not be considered and no correspondence will be entertained in this matter.
- C) Rates must be quoted in Indian rupees and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 1 year minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.

H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.

I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:

- The firm shall have valid GST/Other taxes and IT PAN.
- The firm should not be black listed by any Govt. Agency/Dept.

J) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

K) **Delivery Period** – within 30 days from Purchase order or as specified in Purchase order.

L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.

M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the GMC & SSH, Nagpur.

N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of GMC & SSH, Nagpur with regards to the interpretation of "Terms & Conditions" of this inquiry, the same shall be referred to The Dean, GMC & SSH, Nagpur whose decision will be final and binding upon the contractor.

O) The Dean, GMC & SSH, Nagpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the GMC & SSH, Nagpur will be final in this regard.

P) GMC & SSH, Nagpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the GMC & SSH, Nagpur will be final in this regard.

2. Special Terms & Conditions:

A) Bidder must quote the product as per specification provided in Annexure 1.

B) Catalog must be attached with quotation for technical evaluation.

C) The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the GMC & SSH, Nagpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.



Dean,

**Government Medical College &
Superspeciality Hospital, Nagpur.**

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1
TECHNICAL BID

Sr. No.	Name Of Item	Specification	Qty.
01	Laminar Air Flow	<p>1 Floor model, horizontal flow, well lighted work surface, low vibration and noise, easy to manoeuvre due to castor wheel provision. Over all dimension of work space of approx. 1200 mmx600 mm.600 mm</p> <p>2 Cabinet stainless steel sheet of 20 swg lining</p> <p>3 Front Panels- Removable transparent scratch resistance sheet of approx. 6 mm thickness</p> <p>4 Side Panels- Fixed transparent scratch resistance sheet of approx. 6 mm thickness</p> <p>5 Work Table- Stainless steel of 20 SWG lining</p> <p>6 Pre filters- Filtration efficiency of 98% for all types of particles of sizes of 8 micron and larger.</p> <p>7 Hepafilters (fine Filters)- Filtration efficiency 99.9% For all types of particles fo sies 0.3 micron and larger. Housed in a frame with leak Proof gaskets.</p> <p>8 Motor Blower Dynamically balanced and specially constructed to suit low noise and vibration with adjustable peed. Motor shall conform to ISS or any international specifications.</p> <p>9 Air Velocity- Should not be more than 100 fpm over the work area</p> <p>10 Lighting- Flurescent tube lights with diffuser acrylic to get 120 decalux on work surface</p> <p>11 Ultra Violet light source- Shall be provided</p> <p>12 Power Supply- 20/240 volts, 50 cycles, single phase. The equipment shall be provided with both 5 Amp and plug units inside the cabinet along with a line voltage connector of appropriate rating. Installation, commissioning and trial run will be the responsibility of the supplier</p> <p>13 Line Voltage Corrector- Copper wound single phase automatic line voltage corrector conforming to I:9815/89 with latest amendment fitted with a voltmeter and switch to indicate output/ input voltage as under capacity rating KVA As per the requirement of the equipment Input Voltage- 160 to 260 volts, 50 cycles Output voltage- 220 to 240 volts adjustable The equipment should be supplied with 220 meter cord as input and Fitted with plugs of appropriate rating (15 Amp) Make of the line voltage corrector shall be indicated Manufacturers shall submit the manufacturer's test certificate of each LVC as per IS to be supplied along with the equipment, however type test certificate for4 one number of LVC shall be submitted at the time of inspection</p> <p>14 Manometer-Should be provided with appropriate digital or analog Manometer to measure the air pressure.</p> <p>15 Technical Literature- The firm shall positively submit printed illustrated technical literature/ leaflet including the model quoted by them If quoted model is modified version of their any standard model that also be indicated in the offer.</p> <p>Additional Requirements- All equipment should specify qualifications for design, installation, operation and performance Validation and calibration reports should have traceability to applicable national and international standards. Complete with comprehensive set of spare parts and a suitable capacity voltage stabilize and suitable UPS with maintenance free batteries for minimum 1 hour back-up for each equipment should be supplied with the system. Warranty for 2 tears and CMC/AMC for 8 years with spare parts availability. The made, rating, model, description, specifications, price quantity of each item should be furnished separately. Necessary catalogues, technical write up in English, should be attached with the offer both in hard and electronic copies. Performance, efficiency, other factors as applicable should be furnished. Demonstration and continued comprehensive training for lab staff and support services till familiarity with the system. Should provide electronic and hard copies of user manual (English) services Manual (English) and complete construction details with respect to material specification, thickness finish etc. Should provide a set for equipments for calibration (eg thermometer)and routine Preventive maintenance as per manufacturer documentation in service/ technical manual. Should provide log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.</p>	01

Note:-

Check List.

The following documents are required along with Technical Bid.

- 1) Please mention that the bidder is Manufacturer/Distributor/Dealer/Trader/Supplier relevant Document should be attached.
- 2) Supplier must submit Manufacturer Authorization Certificate.
- 3) Supplier Firm/Company Registration Certificate.
- 4) Copy of PAN Card & GST registration certificate.
- 5) Certificate of USFDA/ISO/CE/DGGI/FDA or any quality certificate.
- 6) Warranty & CMC – Two year warranty & Eight year CMC.
- 7) The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the GMC & Super Speciality Hospital, Nagpur if required. The expenditure incurred for demonstrating the items will be borne by the supplier.
- 8) Relevant brochure/Catalogue pertaining to the item quoted with full specification.
- 9) Certificate of never being black listed by any Organization.
- 10) Have your previously supplied these items to government hospital/Organization? If yes attach the relevant proof.
- 11) Certificate on letter head that you have not quoted the price higher than previously supplied to any government hospital/organization in recent past.
- 12) Performance Security Deposit of 3% of contracted value must be deposited.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BID FORM

To,
The Dean,
GMC & Super Speciality Hospital,
Nagpur.

Dear Sir,

1. I/We _____ Submitted the quotation for Enquiry No.
QUOTATION FOR LAMINOR AIR FLOW AT GMC & SSH, NAGPUR AGAINST ORDER NO.
GMC&SSH/PS/BLOOD BANK/ QT/ /2022 Dt: - / /2022 DUE ON DT: - / /2022
04:00 PM For Laminor Air Flow at GMC & Super Speciality Hospital, Nagpur.

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

Sr. No.	Name of Machine	Qty.	Brand /Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/Unit Inclusive of GST (INR)	MRP
01	Laminor Air Flow	01 Nos.					

Note:-

- 1. The Bidder must quote single Make.**
- The bidder must quoted their quotation only in above said format on the letter of firm otherwise Quotation will be **REJECTED**.
- Catalog must be attached with quotation for technical evaluation.
- The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the GMC & Super Speciality Hospital, Nagpur if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name: - _____

Bank Account No.: _____

IFSC Code: - _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____