

Quotation UPC  
Immediate

Ref NO./Col/Pur/Quot/ 2157/2024  
Office of the Dean,  
Govt. Medical College, Nagpur.  
Date :- 29 / 01 / 2024

To,

M/s -----  
-----  
-----


Sub.:- Submission of e-quotation for the Purchase of Microscope for Pathology Dept at Govt. Medical College, Nagpur.

The undersigned invited sealed quotation for the below on reverse as per enclosed statement for the use of Medical college, Nagpur. on the following terms and conditions...

- 1) The prices quoted should be for delivery at college premises for local dealers and for Nagpur for outsiders.
- 2) The prices quoted by inclusive of all taxes, duties payable like custom excise GST. The breakup of the taxes should also be shown separately where necessary. The sales tax and registration No should be quoted in your letters, Exemption of taxes, if on AF form etc be separately attached.
- 3) The serial no of the items should not be changed while quoting rates. You may dropped in item if not interested.
- 4) Rates should be quoted strictly for the items specified in the list and for standard quality of goods. In case of alternate offer, the detailed specification. Name of manufacturer or make etc. must invariable be stated, specification other than specified in the schedule may be liable for rejection even though lowest.
- 5) The quotation submitted **rates** will be valid for the **period for six months** only from the date of acceptance.
- 6) Delivery period should be stated specifically like ready stock, two weeks/four weeks etc and should be firm and supply of stores if ordered should be made within the stipulated period Failures are liable for delist from the further offers.
- 7) The quotation received after the due date will not be accepted. The quotation should be submitted in the sealed cover. Unsealed quotation will not be accepted. The last date of receiving the quotation is **dt. 05 / 02 / 2024 at. 04:00 PM.**
- 8) Supply of stores should be made in one installment unless otherwise ordered place meal supply will not be accepted. Payment will be made within 4 to 8 weeks after receipt of full quantity and bill in quadruplicates and only satisfactory report of working etc. or part payment will not be accepted.
- 9) Quotation if asked with samples if not accompanies with sample will be liable for re rejection even they are lowest sample should be sent with a liable attached quoting our ref. No of enquiry & item No. etc.

- 10) The Dean, Govt. Medical College, Nagpur. Does not pledge himself to accept the lowest or any quotation and reserve to himself right of acceptance of any quotation which suits to his requirements.
- 11) Very Important :- In case you are not interested in quoting your rates in reply to this enquiry a line in reply is a must failing are liable for desisting their names for further enquires is from our list and no further request in this matter will be entertained.
- 12) Literature, instructions showing specifications working etc. may also be sent with quotations.
- 13) All the nomenclature descriptions, make , pack size and rates should not be quoted with pencil or ink pen , if found, will be rejected and will not be considered. All details mentioned in the quotations must be printed or computerized typed :-

Sr No	Name of item	Qty
01	Binocular Microscope	100

  
( Dr. Raj Gajbhiye )  
Dean  
Govt. Medical College,  
Nagpur

# Specification for Binocular Microscope

## BINOCULAR MICROSCOPES

### Optical System:

- Infinitely corrected optics par focal, plan achromatic lenses with anti fungal properties.

### Illumination:

- Built in transmitted Koehler illumination.
- 6 V, 20 to 30 W halogen bulb
- 220-240V 0.85/0.45A 50Hz

### Focusing

- Stage height movement by roller guide (rock & pinion)
- Upper limit stopper
- Tension adjustable on coarse focus adjustment knob

### Revolving nosepiece

- Quintuple with inward tilt

### Observation tube:

- Tube inclination – 30 -45 0
- Interpupillary distance adjustment range – minimum 50 to 70 mm

### Stage

- Movement range – 76 mm X - direction X 50mm Y - direction
- Rectangular scratch resistant stage with right hand control with double slide holder and vernier calipers on X Y axis.

### Condenser

- Type – Abbe condenser
- N.A. – 1.2 dry type
- Aperture iris diaphragm - built – in

### Objectives, Plan Achromat 4x, 10x, ~~20x~~, 40x & 100x

Minimum working distance for 100X should be 0.13 to 0.2 mm Eyepiece

- 10X with F.N 20

All the necessary adapters and power cords should be provided for functioning of microscope.

### Scope of Supply

- 10 Halogen bulb should be provided with the microscope.

  
Professor & Head  
Department of Pathology  
Govt. Medical College, Nagpur