

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपूर

GOVERNMENT MEDICAL COLLEGE & HOSPITAL, NAGPUR

Tele. No. Office:-2743588

FAX NO. 0712-274448

No.Hosp/PS/QT/Enq/

/2022

Dt. 02/11/2022

Quotation

UPC

21165

To,

M/s.....

.....
.....

Subject: - Submission of Quotation for the Supply of Instrument/Articles/Materials to the Govt. Medical College & Hospital, Nagpur.

The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the Govt. Medical College & Hospital, Nagpur on the following terms & conditions.

- 1) The prices quoted should be for delivery to College & Hospital premises for local dealers and F.O.R. Nagpur for outsiders.
- 2) The Price quoted by inclusive of all taxes, duties if payable like customs, excise, CST, ST, and GST. The break-up of the taxes should also be shown separately wherever necessary. The sales tax & registration number should be quoted in your letter. Exemption of taxes, if on A.F. Forms etc.be separately stated.
- 3) **The serial number of the items should not be changed while quoting rates.** You may drop the item if not interested.
- 4) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 5) The quotation submitted will be valid for a period of One (01) year only from the date of Acceptance.
- 6) Delivery period should be stated specifically , like ready stock or two weeks or four weeks etc and should be made available to this store within the stipulated period. Failures are liable for delist from further enquiries.
- 7) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover and in **separate envelopes** labelled clearly as **"BID 1- Technical" and "BID 2 - Commercial"** Unsealed quotations which are not properly sealed will not be accepted. The last or receiving date of the quotation is **Dt. 14 NOV. 2022** upto 3.00 p.m.
- 8) Supply of stores should be made in one instalment unless ordered otherwise. Piecemeal supply will not be accepted. Condition of advance payment through Bank, partpayment etc. will not be accepted.
- 9) The price quoted is inclusive of all taxes, duties if payable like Customs/ Excise/CST/ST the breakup of the taxes should also be shown separately where necessary. The sales tax and registration number should be quoted in your letter. Exemption of taxes on 'AF/ Octroi exemption form etc.is separately stated.

(Contd. On ...2..)

...2...

- 10) Quotations if asked with samples should be accompanied if not accompanied with sample it will be liable for rejection even they are lowest. Samples should be sent with a labels attached quoting our ref.no of enquiry and item no etc.
- 11) The Dean, Govt. Medical College & Hospital, Nagpur does not pledge himself to accept the lowest or any quotations and reserves the right of acceptance or rejection of any quotations which are not as per the specification or requirement .
- 12) Supplier must quote the rates of items only those which he/she will display/demonstrate and supply. If the good supplied are different from the sample displayed or checked, order will be cancelled & no payment will be done.
- 13) Literature, Instruction manual, working leaflet etc may also be sent with quotation.
- 14) Registration/Authorization Certificate, Sale tax, IT Return Certificate, PAN card, GST no Compliance chart ,brochure ,Catalogue, User List submission (wherever applicable) along With quotation is mandatory.
- 15) Technical Specifications - Compliance Chart with variation/deviation
- 16) Bidder should apply in 2 Bid – TECHNICAL BID & COMMERCIAL BID also write clearly on envelop as TECHNICAL BID and COMMERCIAL BID in BOLD Letter. Bid submitted separately in sealed cover with reference to our Quotation No. on Top of Envelope otherwise their quotation not to be accepted.
- (a) First Bid (Technical Bid). Should including Registration/Authorization Certificate, Sale tax, IT Return Certificate, PAN card, GST no Compliance chart as per specification, Brochure of the quotes Material as per specification write clearly an envelope as TECHNICAL BID.
- (b) Second Bid (commercial Bid). Should include items rate including all Taxes only.
- 17) While submitting the bills of good supplied, supplier must calculate unit price and final unit price should reflect the same rate as per order given by the Institute. GST & other taxes should not be calculated on total prices.
- 18) After Opening the technical BD qualified Bidder will have to give demonstration of equipment as well as show their quoted items to HOD for technical approval as a Technical Committee otherwise Bidder names not to be included in commercial bid.



DEAN

Govt. Medical College & Hospital,
Nagpur

Details as Below:

Sr.No	Name of Item with specification	Qty	Rate Incl.GST
1	Electric Orthotic Oven (Spec Attach)	1Nos	
2	Prosthetic Oven (Spec Attach)	1Nos	

Note: Rate are Including GST and all taxes.

Requirement of Orthotic oven

- 1) Double walled construction - Inside chamber made of stainless steel & outside made of mild steel.
painted & stone enamel.
- 2) The temperature is controlled by Digital indicator cum controller
- 3) Heating is uniform all around through circulating fans \Rightarrow on top
- 4) supplied & adjustable trays.
- 5) Toughened glass window on the door.
- 6) A digital timer along & an alarm system
- 7) Interior dimension LXBXH - 38" x 30" x 24"
- 8) Temperature Ranges to 300 degree C.
- 9) power Requirement - V/HZ/KW - 220-230/50-60/5.5.
single phase current - / ~~single~~ 3 phase

Yamraj

Requirement of prothetic oven

- 1) Double walled construction - inside chamber made of stainless steel & outside made of mild steel painted & stove enamel.
- 2) The temperature is controlled by Digital indicator cum controller
- 3) Heating is uniform all around through circulating fans.
- 4) Supplied & adjustable trays.
- 5) Toughened glass window on the door.
- 6) A Digital timer along & an alarm system
- 7) inside chamber size - 24" x 30" x 20" (L x B x H)
- 8) Temperature ranges is upto 300 degree c.
- 9) Dimension 113 x 67 x 100 cm (L x B x H)

EIL-SM-25

Power Input	phase	volts	weight
5.5W	single/3 phase	220-240V	65kg.

