

Quotation UPC

Immediate

Ref NO./Col/Pur/Quot/ 16669 /2023

Office of the Dean,

Govt. Medical College, Nagpur.

Date :- 06 / 09 / 2023

To,

M/s -----


Sub.:- Submission of e-quotation for the Purchase of items for Pathology Dept at Govt. Medical College Nagpur.

The undersigned invited sealed quotation for the below on reverse as per enclosed statement for the use of Medical college, Nagpur. on the following terms and conditions...

- 1) The prices quoted should be for delivery at college premises for local dealers and for Nagpur for outsiders.
- 2) The prices quoted by inclusive of all taxes, duties payable like custom excise GST. The breakup of the taxes should also be shown separately where necessary. The sales tax and registration No should be quoted in your letters, Exemption of taxes, if on AF form etc be separately attached.
- 3) The serial no of the items should not be changed while quoting rates. You may dropped in item if not interested.
- 4) Rates should be quoted strictly for the items specified in the list and for standard quality of goods. Incase of alternate offer, the detailed specification. Name of manufacturer or make etc. must invariable be stated, specification other than specified in the schedule may be liable for rejection aven thought lowest.
- 5) The quotation submitted **rates** will be valid for the **period for one year** only from the date of acceptance.
- 6) Delivery period should be stated specifically like ready stock, two weeks/four weeks etc and should be firm and supply of stores if ordered should be made with in the stipulated period Failures are liable for delist from the further offers.
- 7) The quotation received after the due date will not be accepted. The quotation should be submitted in the sealed cover. Unsealed quotation will not be accepted. The last date of receiving the quotation is **dt. 13 / 09 / 2023 at. 05:00 PM.**
- 8) Supply of stores should be made in one installment unless otherwise ordered place meal supply will not be accepted. Payment will be made within 4 to 8 weeks after receipt of full quantity and bill in quadruplicates and only satisfactory report of working etc. or part payment will not be accepted.
- 9) Quotation if asked with samples if not accompanies with sample will be liable for re rejection even they are lowest sample should be sent with a liable attached quoting our ref. No of enquiry & item No. etc.

- 10) The Dean, Govt. Medical College, Nagpur. Does not pledge himself to accept the lowest or any quotation and reserve to himself right of acceptance of any quotation which suits to his requirements.
- 11) Very Important :- In case you are not interested in quoting your rates in reply to this enquiry a line in reply is a must failing are liable for desisting their names for further enquires is from our list and no further request in this matter will be entertained.
- 12) Literature, instructions showing specifications working etc. may also be sent with quotations.
- 13) All the nomenclature descriptions, make , pack size and rates should not be quoted with pencil or ink pen , if found, will be rejected and will not be considered. All details mentioned in the quotations must be printed or computerized typed :-

Sr No	Name of item	Qty
01	Electrophoresis , Cellulose acetate haemoglobin electrophoresis sysytem	01
02	Water Bath	01
03	Hot Plate	02
04	Distilled water plant	01


(Dr. Raj Gajbhiye)
Dean,
Govt. Medical College,
Nagpur

Water Bath Technical Specifications

Temp – 0 c to 60 c

Capacity 24 lit +/- = 4

Inner Material – Stainless Steel

Dimensions -45 X 30 X 17 cm

No. of Racks – 4 No.

Outer Material – Steel

Power – 2 kw

Voltage – 220 – 230 v AC

LABORATORY HOT PLATE

Lab grade heating plates for warm applications

Steribox offers you wide range of laboratory hot plates, available in different sizes and temperature ranges. We make laboratory hot plates in both rectangular and round shapes with surface temperature range up to 350°C and optional 500°C. Heating surface is made of either MS, Stainless Steel or Cast Iron. Temperature is controlled through energy regulator or digital PID controller which efficiently controls temperature and also displays set value and process value. Other cabinet is made of powder coated mild steel. These lab grade heating plates are light weight and durable, and can be bought at reasonable price in India.

TECHNICAL SPECIFICATIONS

Plate design	Rectangular and Round
Temperature	Up to 300°C
Heating plate surface	Stainless Steel 304 / Mild Steel / Cast Iron
Cabinet MOC	Powder coated MS
Controller	Thermostatic controller
Power supply	220 Volts 50Hz
Optional	<ul style="list-style-type: none"> ■ Digital temperature controller ■ Temperature up to 500°C

STANDARD SIZES / DIMENSIONS

Model	Plate Dimensions (L*W)	Watt
STXHP1012	250 x 300 mm / 10*12 Inch	1.2 KW
STXHP1610	250 x 400 mm / 10*16 Inch	1.5 KW
STXHP1612	300 x 455 mm / 12*18 Inch	2.0 KW
STXHP1616	455 x 455 mm / 18*18 Inch	3.0 KW
STXHP2416	455 x 600 mm / 18*24 Inch	3.0 KW
STXHP2424	600 x 600 mm / 24*24 Inch	3.0 KW
STXHP2436	600 x 900 mm / 24*36 Inch	3.0 KW
STXRHP8	8 (Dia.)	1.5 KW
STXRHP12	12 (Dia.)	1.5 KW

Specification

Distillation water plant

- 1} table top model
- 2} Type :- Single distillation
- 3} Capacity :- 05 liter
- 4} Power :- 01 kw
- 5} Material :- Borosilicate glass
- 5 } Flask Colour:- Transparent
- 6 } Power Source :- electric
- 7} Temp. Deg. Celsius -25- 200 c

Automatic electrically heated glass distillation unit heater embedded in glass tube.

An automatic cut off device attached with apparatus with proper (leak proof) connection to tap & container.

Specification

Water Bath

Temp - 0 c to 60 c

Capacity 24 lit +/-4

Inner Material - stainless steel

Dimensions -45 X 30 X 17 cm

No. of Racks -4 No.

Outer material - Steel

Power - 2 kw

Voltage -220 - 230v Ac

Electrophoresis Specification

Power Supply 4 Output terminals output

- 1) Constant Voltage :- 10 to 300 V DC.
- 2) Constant Current :- 4 to 400 mA.